

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090624\
 Data File : P0106267.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2024 21:01
 Operator : YP/AJ
 Sample : P3851-11
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC271246-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/09/2024
 Supervised By :Ankita Jodhani 09/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 07:22:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.453	3.718	108.0E6	38942566	11.837	14.414m
2) SA Decachlor...	10.195	8.746	99052148	36920868	18.252	16.913m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090624\
Data File : PO106267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2024 21:01
Operator : YP/AJ
Sample : P3851-11
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC271246-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/09/2024
Supervised By :Ankita Jodhani 09/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 07:22:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 06:50:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.453 min
Delta R.T.: -0.001 min
Response: 107988373
Conc: 11.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC271246-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/09/2024
Supervised By :Ankita Jodhani 09/09/2024




#1 Tetrachloro-m-xylene

R.T.: 3.718 min
Delta R.T.: -0.004 min
Response: 38942566
Conc: 14.41 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.195 min
Delta R.T.: 0.000 min
Response: 99052148
Conc: 18.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.746 min
Delta R.T.: -0.011 min
Response: 36920868
Conc: 16.91 ng/ml m