

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090624\
 Data File : P0106308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2024 12:26
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 09 00:59:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 27 06:50:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.716	496.8E6	185.0E6	54.449	68.483 #
2) SA Decachlor...	10.196	8.746	206.6E6	86705863	38.079	39.718
Target Compounds						
26) L6 AR-1254-1	6.471	5.610	131.4E6	77332499	425.726	573.160 #
27) L6 AR-1254-2	6.689	5.757	187.3E6	65753221	406.608	541.755 #
28) L6 AR-1254-3	7.054	6.162	175.9E6	96672580	372.339	501.521 #
29) L6 AR-1254-4	7.339	6.390	115.6E6	55602135	331.883	458.488 #
30) L6 AR-1254-5	7.757	6.808	122.3E6	75929963	335.001	429.448 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090624\
Data File : PO106308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2024 12:26
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 09 00:59:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 27 06:50:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.452 min
Delta R.T.: -0.002 min
Response: 496751161
Conc: 54.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.716 min
Delta R.T.: -0.005 min
Response: 185024352
Conc: 68.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.196 min
Delta R.T.: 0.000 min
Response: 206649053
Conc: 38.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.746 min
Delta R.T.: -0.011 min
Response: 86705863
Conc: 39.72 ng/ml



#26 AR-1254-1


R.T.: 6.471 min
Delta R.T.: 0.000 min
Response: 131424738
Conc: 425.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.610 min
Delta R.T.: -0.007 min
Response: 77332499
Conc: 573.16 ng/ml




#27 AR-1254-2

R.T.: 6.689 min
Delta R.T.: 0.000 min
Response: 187334026
Conc: 406.61 ng/ml



#27 AR-1254-2


R.T.: 5.757 min
Delta R.T.: -0.007 min
Response: 65753221
Conc: 541.75 ng/ml



#28 AR-1254-3


R.T.: 7.054 min
Delta R.T.: 0.001 min
Response: 175898459
Conc: 372.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.162 min
Delta R.T.: -0.007 min
Response: 96672580
Conc: 501.52 ng/ml




#29 AR-1254-4

R.T.: 7.339 min
Delta R.T.: 0.001 min
Response: 115610767
Conc: 331.88 ng/ml



#29 AR-1254-4


R.T.: 6.390 min
Delta R.T.: -0.007 min
Response: 55602135
Conc: 458.49 ng/ml



#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: 0.001 min
Response: 122327048
Conc: 335.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.808 min
Delta R.T.: -0.008 min
Response: 75929963
Conc: 429.45 ng/ml