

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090722\
 Data File : P0089368.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2022 13:43
 Operator : YP/AJ
 Sample : N4485-05DL 2000X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CONCRETE-COMP-2DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 14:25:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
26)	L6	AR-1254-1	6.465	5.459	91992648 48990049	799.815 828.071
27)	L6	AR-1254-2	6.684	5.606	143.6E6 47862662	843.899 938.278
28)	L6	AR-1254-3	7.053	6.008	153.1E6 81099872	881.629 1026.920
29)	L6	AR-1254-4	7.338	6.238	109.0E6 49185671	872.824 992.253
30)	L6	AR-1254-5	7.760	6.654	143.2E6 89622072	1112.029 1325.551



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090722\
Data File : PO089368.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2022 13:43
Operator : YP/AJ
Sample : N4485-05DL 2000X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CONCRETE-COMP-2DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 14:25:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.430 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
CONCRETE-COMP-2DL




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.589 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.257 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.572 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.465 min
Delta R.T.: -0.002 min
Response: 91992648
Conc: 799.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-COMP-2DL




#26 AR-1254-1

R.T.: 5.459 min
Delta R.T.: -0.006 min
Response: 48990049
Conc: 828.07 ng/ml




#27 AR-1254-2

R.T.: 6.684 min
Delta R.T.: -0.002 min
Response: 143602341
Conc: 843.90 ng/ml



#27 AR-1254-2


R.T.: 5.606 min
Delta R.T.: -0.006 min
Response: 47862662
Conc: 938.28 ng/ml



#28 AR-1254-3


R.T.: 7.053 min
Delta R.T.: -0.002 min
Response: 153119559
Conc: 881.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-COMP-2DL




#28 AR-1254-3

R.T.: 6.008 min
Delta R.T.: -0.006 min
Response: 81099872
Conc: 1026.92 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: -0.003 min
Response: 108951126
Conc: 872.82 ng/ml



#29 AR-1254-4


R.T.: 6.238 min
Delta R.T.: -0.006 min
Response: 49185671
Conc: 992.25 ng/ml



#30 AR-1254-5

R.T.: 7.760 min
Delta R.T.: -0.002 min
Response: 143231888
Conc: 1112.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-COMP-2DL



#30 AR-1254-5

R.T.: 6.654 min
Delta R.T.: -0.006 min
Response: 89622072
Conc: 1325.55 ng/ml