

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090723\
 Data File : P0097811.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2023 00:15
 Operator : YP/AJ
 Sample : PB155370BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB155370BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 02:03:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.442	3.622	91899694	21872544	23.067	22.839
2) SA Decachlor...	10.289	8.659	46147939	20211423	24.698	23.656
Target Compounds						
3) L1 AR-1016-1	5.619	4.713	5101832	1415993	47.248	44.024
4) L1 AR-1016-2	5.642	4.732	7309946	1956298	45.171	43.066
5) L1 AR-1016-3	5.705	4.909	4338479	1112782	43.571	45.705
6) L1 AR-1016-4	5.804	4.951	3328235	965515	42.207	45.957
7) L1 AR-1016-5	6.101	5.166	3376494	1282285	41.129	46.138
31) L7 AR-1260-1	7.239	6.204	5904193	2380021	41.779	44.733
32) L7 AR-1260-2	7.497	6.393	6439154	3094615	42.599	49.420
33) L7 AR-1260-3	7.861	6.547	4485051	2554256	39.593	44.295
34) L7 AR-1260-4	8.089	7.021	6362270	2012470	53.521	42.900
35) L7 AR-1260-5	8.417	7.263	8825302	4353268	43.109	43.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090723\
Data File : PO097811.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2023 00:15
Operator : YP/AJ
Sample : PB155370BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB155370BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 02:03:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

