

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090723\
 Data File : P0097817.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2023 02:11
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 03:14:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.441	3.622	198.8E6	50278774	49.906	52.500
2) SA Decachlor...	10.288	8.657	94004475	42393260	50.311	49.619
Target Compounds						
16) L4 AR-1242-1	5.618	4.712	47774702	14884725	512.923	530.401
17) L4 AR-1242-2	5.641	4.731	70136601	20476731	499.909	515.935
18) L4 AR-1242-3	5.704	4.907	43473303	11375243	502.316	538.295
19) L4 AR-1242-4	5.803	4.992	34059971	11925275	502.205	520.333
20) L4 AR-1242-5	6.546	5.517	35152074	13929741	499.550	534.705



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090723\
Data File : PO097817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2023 02:11
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 03:14:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.441 min
Delta R.T.: -0.002 min
Response: 198830084
Conc: 49.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.001 min
Response: 50278774
Conc: 52.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.288 min
Delta R.T.: -0.007 min
Response: 94004475
Conc: 50.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.657 min
Delta R.T.: -0.007 min
Response: 42393260
Conc: 49.62 ng/ml



#16 AR-1242-1


R.T.: 5.618 min
Delta R.T.: -0.003 min
Response: 47774702
Conc: 512.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.712 min
Delta R.T.: -0.002 min
Response: 14884725
Conc: 530.40 ng/ml




#17 AR-1242-2

R.T.: 5.641 min
Delta R.T.: -0.003 min
Response: 70136601
Conc: 499.91 ng/ml



#17 AR-1242-2


R.T.: 4.731 min
Delta R.T.: -0.003 min
Response: 20476731
Conc: 515.94 ng/ml



#18 AR-1242-3


R.T.: 5.704 min
Delta R.T.: -0.003 min
Response: 43473303
Conc: 502.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.907 min
Delta R.T.: -0.003 min
Response: 11375243
Conc: 538.30 ng/ml




#19 AR-1242-4

R.T.: 5.803 min
Delta R.T.: -0.003 min
Response: 34059971
Conc: 502.20 ng/ml



#19 AR-1242-4


R.T.: 4.992 min
Delta R.T.: -0.003 min
Response: 11925275
Conc: 520.33 ng/ml



#20 AR-1242-5

R.T.: 6.546 min
Delta R.T.: -0.003 min
Response: 35152074
Conc: 499.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.517 min
Delta R.T.: -0.003 min
Response: 13929741
Conc: 534.70 ng/ml