

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090822\
 Data File : P0089400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2022 12:32
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 16:56:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0082922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 29 17:48:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.583	213.6E6	54112527	56.962	56.060
2) SA Decachlor...	10.263	8.568	116.0E6	51628442	50.365	58.384
Target Compounds						
3) L1 AR-1016-1	5.603	4.658	60425722	19031658	537.424	527.000
4) L1 AR-1016-2	5.626	4.676	86409152	26710192	548.287	541.028
5) L1 AR-1016-3	5.688	4.851	53570827	14324897	549.287	538.588
6) L1 AR-1016-4	5.788	4.893	42352976	11339863	538.249	526.865
7) L1 AR-1016-5	6.084	5.105	40870134	14400361	532.554	493.538
31) L7 AR-1260-1	7.218	6.136	69441510	27927951	515.698	497.795
32) L7 AR-1260-2	7.475	6.325	78255966	33201007	512.772	488.594
33) L7 AR-1260-3	7.837	6.477	58619872	30144731	524.604	491.326
34) L7 AR-1260-4	8.065	6.949	61451111	24764328	497.785	491.430
35) L7 AR-1260-5	8.392	7.192	117.9E6	58544613	494.048	511.683



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090822\
Data File : PO089400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2022 12:32
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 16:56:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO082922.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 29 17:48:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.427 min
Delta R.T.: -0.003 min
Response: 213586190
Conc: 56.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.583 min
Delta R.T.: -0.006 min
Response: 54112527
Conc: 56.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.263 min
Delta R.T.: 0.007 min
Response: 115956093
Conc: 50.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.568 min
Delta R.T.: -0.005 min
Response: 51628442
Conc: 58.38 ng/ml



#3 AR-1016-1


R.T.: 5.603 min
Delta R.T.: -0.001 min
Response: 60425722
Conc: 537.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.658 min
Delta R.T.: -0.007 min
Response: 19031658
Conc: 527.00 ng/ml




#4 AR-1016-2

R.T.: 5.626 min
Delta R.T.: -0.001 min
Response: 86409152
Conc: 548.29 ng/ml



#4 AR-1016-2


R.T.: 4.676 min
Delta R.T.: -0.008 min
Response: 26710192
Conc: 541.03 ng/ml



#5 AR-1016-3


R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 53570827
Conc: 549.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.851 min
Delta R.T.: -0.008 min
Response: 14324897
Conc: 538.59 ng/ml




#6 AR-1016-4

R.T.: 5.788 min
Delta R.T.: -0.002 min
Response: 42352976
Conc: 538.25 ng/ml



#6 AR-1016-4


R.T.: 4.893 min
Delta R.T.: -0.007 min
Response: 11339863
Conc: 526.87 ng/ml



#7 AR-1016-5


R.T.: 6.084 min
Delta R.T.: 0.000 min
Response: 40870134
Conc: 532.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.105 min
Delta R.T.: -0.008 min
Response: 14400361
Conc: 493.54 ng/ml




#31 AR-1260-1

R.T.: 7.218 min
Delta R.T.: 0.001 min
Response: 69441510
Conc: 515.70 ng/ml



#31 AR-1260-1


R.T.: 6.136 min
Delta R.T.: -0.007 min
Response: 27927951
Conc: 497.79 ng/ml



#32 AR-1260-2


R.T.: 7.475 min
Delta R.T.: 0.000 min
Response: 78255966
Conc: 512.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.325 min
Delta R.T.: -0.006 min
Response: 33201007
Conc: 488.59 ng/ml




#33 AR-1260-3

R.T.: 7.837 min
Delta R.T.: 0.000 min
Response: 58619872
Conc: 524.60 ng/ml



#33 AR-1260-3


R.T.: 6.477 min
Delta R.T.: -0.007 min
Response: 30144731
Conc: 491.33 ng/ml



#34 AR-1260-4


R.T.: 8.065 min
Delta R.T.: 0.002 min
Response: 61451111
Conc: 497.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.949 min
Delta R.T.: -0.006 min
Response: 24764328
Conc: 491.43 ng/ml



#35 AR-1260-5

R.T.: 8.392 min
Delta R.T.: 0.003 min
Response: 117859311
Conc: 494.05 ng/ml



#35 AR-1260-5

R.T.: 7.192 min
Delta R.T.: -0.005 min
Response: 58544613
Conc: 511.68 ng/ml