

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090919\
 Data File : P0060567.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Sep 2019 18:24
 Operator : SM/AJ
 Sample : K4759-03MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 07:21:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Sep 08 02:12:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.368	3.534	796236	672889	19.190	21.908
2)	SA Decachlor...	10.030	8.423	903244	711811	17.123	19.229
Target Compounds							
3)	L1 AR-1016-1	5.532	4.591	376280	309957	231.267	246.831
4)	L1 AR-1016-2	5.555	4.609	539001	428887	229.848	245.456
5)	L1 AR-1016-3	5.616	4.780	329148	239289	226.100	251.644
6)	L1 AR-1016-4	5.715	4.822	266245	198260	216.558	240.559
7)	L1 AR-1016-5	6.008	5.029	278112	258136	225.471	242.951
8)	L2 AR-1221-1	4.571	3.743	25555	25542	56.278	65.571
9)	L2 AR-1221-2	4.665	3.826	40574	33945	116.229	113.490
10)	L2 AR-1221-3	4.740	3.901	156779	141343	137.268	154.468
11)	L3 AR-1232-1	4.740	3.901	156779	141343	126.894	144.572
12)	L3 AR-1232-2	5.266	4.609	245019	428887	359.369	406.990
13)	L3 AR-1232-3	5.555	4.780	539001	239289	384.431	422.558
14)	L3 AR-1232-4	5.715	4.862	266245	243722	364.112	448.151
15)	L3 AR-1232-5	5.805	5.029	246511	258136	471.512	432.651
16)	L4 AR-1242-1	5.532	4.591	376280	309957	312.335	337.156
17)	L4 AR-1242-2	5.555	4.609	539001	428887	311.660	334.349
18)	L4 AR-1242-3	5.616	4.780	329148	239289	298.941	335.133
19)	L4 AR-1242-4	5.715	4.862	266245	243722	295.752	325.058
20)	L4 AR-1242-5	6.449	5.373	39935	211544	36.637	216.654 #
21)	L5 AR-1248-1	5.532	4.591	376280	309957	410.066	428.799
22)	L5 AR-1248-2	5.805	4.822	246511	198260	184.004	202.446
23)	L5 AR-1248-3	6.008	4.862	278112	243722	185.560	238.808 #
24)	L5 AR-1248-4	6.410	5.029	44147	258136	25.100	212.300 #
25)	L5 AR-1248-5	6.449	5.411	39935	34739	23.872	28.553
26)	L6 AR-1254-1	6.383	5.373	224417	211544	123.054	111.494
27)	L6 AR-1254-2	6.600	5.517	258844	210256	89.313	123.031 #
28)	L6 AR-1254-3	6.966	5.928	142477	391119	44.676	139.457 #
29)	L6 AR-1254-4	7.250	6.136	91445	47090	38.081	26.358 #
30)	L6 AR-1254-5	7.666	6.547	887427	741296	326.227	282.505
31)	L7 AR-1260-1	7.128	6.040	612373	524978	244.978	255.288
32)	L7 AR-1260-2	7.384	6.226	732176	625286	235.621	249.737
33)	L7 AR-1260-3	7.742	6.376	475925	586380	190.263	252.391 #
34)	L7 AR-1260-4	7.966	6.842	593613	445699	200.493	219.442
35)	L7 AR-1260-5	8.280	7.081	1028239	955562	183.064	208.601
36)	L8 AR-1262-1	7.742	6.586	475925	469964	141.717	176.755
37)	L8 AR-1262-2	8.280	7.081	1028239	955562	175.624	210.493
38)	L8 AR-1262-3	8.598	7.361	663783	196368	165.880	103.591 #
39)	L8 AR-1262-4	8.654	7.423	361555	692430	119.058	201.918 #
40)	L8 AR-1262-5	9.306	7.911	230145	204489	114.736	130.785
41)	L9 AR-1268-1	8.598	7.361	663783	196368	96.154	38.664 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090919\
 Data File : P0060567.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Sep 2019 18:24
 Operator : SM/AJ
 Sample : K4759-03MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 07:21:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Sep 08 02:12:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.654	7.423	361555	692430	56.220	149.068 #
43) L9	AR-1268-3	8.894	7.630	22418	20082	4.229	5.107
44) L9	AR-1268-4	9.306	7.911	230145	204489	103.738	120.265
45) L9	AR-1268-5	9.706	8.185	78813	68321	4.810	6.098 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0090919\
Data File : P0060567.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Sep 2019 18:24
Operator : SM/AJ
Sample : K4759-03MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 07:21:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0090719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Sep 08 02:12:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.368 min
Delta R.T.: 0.000 min
Response: 796236
Conc: 19.19 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.534 min
Delta R.T.: -0.002 min
Response: 672889
Conc: 21.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.030 min
Delta R.T.: -0.008 min
Response: 903244
Conc: 17.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.423 min
Delta R.T.: -0.008 min
Response: 711811
Conc: 19.23 ng/ml



#3 AR-1016-1


R.T.: 5.532 min
Delta R.T.: -0.002 min
Response: 376280
Conc: 231.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.591 min
Delta R.T.: -0.004 min
Response: 309957
Conc: 246.83 ng/ml




#4 AR-1016-2

R.T.: 5.555 min
Delta R.T.: -0.001 min
Response: 539001
Conc: 229.85 ng/ml



#4 AR-1016-2


R.T.: 4.609 min
Delta R.T.: -0.003 min
Response: 428887
Conc: 245.46 ng/ml



#5 AR-1016-3


R.T.: 5.616 min
Delta R.T.: -0.002 min
Response: 329148
Conc: 226.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.780 min
Delta R.T.: -0.004 min
Response: 239289
Conc: 251.64 ng/ml




#6 AR-1016-4

R.T.: 5.715 min
Delta R.T.: -0.002 min
Response: 266245
Conc: 216.56 ng/ml



#6 AR-1016-4


R.T.: 4.822 min
Delta R.T.: -0.004 min
Response: 198260
Conc: 240.56 ng/ml



#7 AR-1016-5


R.T.: 6.008 min
Delta R.T.: -0.002 min
Response: 278112
Conc: 225.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.029 min
Delta R.T.: -0.004 min
Response: 258136
Conc: 242.95 ng/ml




#8 AR-1221-1

R.T.: 4.571 min
Delta R.T.: -0.006 min
Response: 25555
Conc: 56.28 ng/ml



#8 AR-1221-1


R.T.: 3.743 min
Delta R.T.: -0.004 min
Response: 25542
Conc: 65.57 ng/ml



#9 AR-1221-2


R.T.: 4.665 min
Delta R.T.: 0.001 min
Response: 40574
Conc: 116.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.826 min
Delta R.T.: -0.004 min
Response: 33945
Conc: 113.49 ng/ml




#10 AR-1221-3

R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 156779
Conc: 137.27 ng/ml



#10 AR-1221-3


R.T.: 3.901 min
Delta R.T.: -0.003 min
Response: 141343
Conc: 154.47 ng/ml



#11 AR-1232-1


R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 156779
Conc: 126.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.901 min
Delta R.T.: -0.003 min
Response: 141343
Conc: 144.57 ng/ml




#12 AR-1232-2

R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 245019
Conc: 359.37 ng/ml



#12 AR-1232-2


R.T.: 4.609 min
Delta R.T.: -0.003 min
Response: 428887
Conc: 406.99 ng/ml



#13 AR-1232-3


R.T.: 5.555 min
Delta R.T.: 0.000 min
Response: 539001
Conc: 384.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.780 min
Delta R.T.: -0.003 min
Response: 239289
Conc: 422.56 ng/ml




#14 AR-1232-4

R.T.: 5.715 min
Delta R.T.: -0.002 min
Response: 266245
Conc: 364.11 ng/ml



#14 AR-1232-4


R.T.: 4.862 min
Delta R.T.: -0.003 min
Response: 243722
Conc: 448.15 ng/ml



#15 AR-1232-5


R.T.: 5.805 min
Delta R.T.: -0.001 min
Response: 246511
Conc: 471.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.029 min
Delta R.T.: -0.003 min
Response: 258136
Conc: 432.65 ng/ml




#16 AR-1242-1

R.T.: 5.532 min
Delta R.T.: -0.003 min
Response: 376280
Conc: 312.33 ng/ml



#16 AR-1242-1


R.T.: 4.591 min
Delta R.T.: -0.003 min
Response: 309957
Conc: 337.16 ng/ml



#17 AR-1242-2


R.T.: 5.555 min
Delta R.T.: -0.001 min
Response: 539001
Conc: 311.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.609 min
Delta R.T.: -0.003 min
Response: 428887
Conc: 334.35 ng/ml




#18 AR-1242-3

R.T.: 5.616 min
Delta R.T.: -0.002 min
Response: 329148
Conc: 298.94 ng/ml



#18 AR-1242-3


R.T.: 4.780 min
Delta R.T.: -0.003 min
Response: 239289
Conc: 335.13 ng/ml



#19 AR-1242-4


R.T.: 5.715 min
Delta R.T.: -0.002 min
Response: 266245
Conc: 295.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.862 min
Delta R.T.: -0.003 min
Response: 243722
Conc: 325.06 ng/ml




#20 AR-1242-5

R.T.: 6.449 min
Delta R.T.: -0.001 min
Response: 39935
Conc: 36.64 ng/ml



#20 AR-1242-5


R.T.: 5.373 min
Delta R.T.: -0.003 min
Response: 211544
Conc: 216.65 ng/ml



#21 AR-1248-1


R.T.: 5.532 min
Delta R.T.: 0.000 min
Response: 376280
Conc: 410.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.591 min
Delta R.T.: -0.002 min
Response: 309957
Conc: 428.80 ng/ml




#22 AR-1248-2

R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 246511
Conc: 184.00 ng/ml



#22 AR-1248-2


R.T.: 4.822 min
Delta R.T.: -0.003 min
Response: 198260
Conc: 202.45 ng/ml



#23 AR-1248-3


R.T.: 6.008 min
Delta R.T.: -0.001 min
Response: 278112
Conc: 185.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.862 min
Delta R.T.: -0.003 min
Response: 243722
Conc: 238.81 ng/ml




#24 AR-1248-4

R.T.: 6.410 min
Delta R.T.: -0.001 min
Response: 44147
Conc: 25.10 ng/ml



#24 AR-1248-4


R.T.: 5.029 min
Delta R.T.: -0.003 min
Response: 258136
Conc: 212.30 ng/ml



#25 AR-1248-5


R.T.: 6.449 min
Delta R.T.: 0.000 min
Response: 39935
Conc: 23.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.411 min
Delta R.T.: -0.004 min
Response: 34739
Conc: 28.55 ng/ml




#26 AR-1254-1

R.T.: 6.383 min
Delta R.T.: -0.001 min
Response: 224417
Conc: 123.05 ng/ml



#26 AR-1254-1


R.T.: 5.373 min
Delta R.T.: -0.004 min
Response: 211544
Conc: 111.49 ng/ml



#27 AR-1254-2


R.T.: 6.600 min
Delta R.T.: -0.003 min
Response: 258844
Conc: 89.31 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.517 min
Delta R.T.: -0.004 min
Response: 210256
Conc: 123.03 ng/ml




#28 AR-1254-3

R.T.: 6.966 min
Delta R.T.: -0.002 min
Response: 142477
Conc: 44.68 ng/ml



#28 AR-1254-3


R.T.: 5.928 min
Delta R.T.: 0.011 min
Response: 391119
Conc: 139.46 ng/ml



#29 AR-1254-4


R.T.: 7.250 min
Delta R.T.: -0.003 min
Response: 91445
Conc: 38.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.136 min
Delta R.T.: -0.005 min
Response: 47090
Conc: 26.36 ng/ml




#30 AR-1254-5

R.T.: 7.666 min
Delta R.T.: -0.004 min
Response: 887427
Conc: 326.23 ng/ml



#30 AR-1254-5


R.T.: 6.547 min
Delta R.T.: -0.005 min
Response: 741296
Conc: 282.51 ng/ml



#31 AR-1260-1


R.T.: 7.128 min
Delta R.T.: -0.002 min
Response: 612373
Conc: 244.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.040 min
Delta R.T.: -0.006 min
Response: 524978
Conc: 255.29 ng/ml




#32 AR-1260-2

R.T.: 7.384 min
Delta R.T.: -0.004 min
Response: 732176
Conc: 235.62 ng/ml



#32 AR-1260-2


R.T.: 6.226 min
Delta R.T.: -0.005 min
Response: 625286
Conc: 249.74 ng/ml



#33 AR-1260-3


R.T.: 7.742 min
Delta R.T.: -0.004 min
Response: 475925
Conc: 190.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.376 min
Delta R.T.: -0.005 min
Response: 586380
Conc: 252.39 ng/ml




#34 AR-1260-4

R.T.: 7.966 min
Delta R.T.: -0.004 min
Response: 593613
Conc: 200.49 ng/ml



#34 AR-1260-4


R.T.: 6.842 min
Delta R.T.: -0.006 min
Response: 445699
Conc: 219.44 ng/ml



#35 AR-1260-5


R.T.: 8.280 min
Delta R.T.: -0.004 min
Response: 1028239
Conc: 183.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.081 min
Delta R.T.: -0.006 min
Response: 955562
Conc: 208.60 ng/ml




#36 AR-1262-1

R.T.: 7.742 min
Delta R.T.: -0.003 min
Response: 475925
Conc: 141.72 ng/ml



#36 AR-1262-1


R.T.: 6.586 min
Delta R.T.: -0.005 min
Response: 469964
Conc: 176.76 ng/ml



#37 AR-1262-2


R.T.: 8.280 min
Delta R.T.: -0.004 min
Response: 1028239
Conc: 175.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.081 min
Delta R.T.: -0.005 min
Response: 955562
Conc: 210.49 ng/ml




#38 AR-1262-3

R.T.: 8.598 min
Delta R.T.: 0.009 min
Response: 663783
Conc: 165.88 ng/ml



#38 AR-1262-3


R.T.: 7.361 min
Delta R.T.: -0.006 min
Response: 196368
Conc: 103.59 ng/ml



#39 AR-1262-4


R.T.: 8.654 min
Delta R.T.: -0.022 min
Response: 361555
Conc: 119.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.423 min
Delta R.T.: -0.007 min
Response: 692430
Conc: 201.92 ng/ml




#40 AR-1262-5

R.T.: 9.306 min
Delta R.T.: -0.006 min
Response: 230145
Conc: 114.74 ng/ml



#40 AR-1262-5


R.T.: 7.911 min
Delta R.T.: -0.006 min
Response: 204489
Conc: 130.78 ng/ml



#41 AR-1268-1


R.T.: 8.598 min
Delta R.T.: 0.013 min
Response: 663783
Conc: 96.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.361 min
Delta R.T.: -0.006 min
Response: 196368
Conc: 38.66 ng/ml




#42 AR-1268-2

R.T.: 8.654 min
Delta R.T.: -0.024 min
Response: 361555
Conc: 56.22 ng/ml



#42 AR-1268-2


R.T.: 7.423 min
Delta R.T.: -0.007 min
Response: 692430
Conc: 149.07 ng/ml



#43 AR-1268-3


R.T.: 8.894 min
Delta R.T.: -0.004 min
Response: 22418
Conc: 4.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.630 min
Delta R.T.: -0.006 min
Response: 20082
Conc: 5.11 ng/ml




#44 AR-1268-4

R.T.: 9.306 min
Delta R.T.: -0.006 min
Response: 230145
Conc: 103.74 ng/ml



#44 AR-1268-4


R.T.: 7.911 min
Delta R.T.: -0.006 min
Response: 204489
Conc: 120.27 ng/ml



#45 AR-1268-5

R.T.: 9.706 min
Delta R.T.: -0.005 min
Response: 78813
Conc: 4.81 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.185 min
Delta R.T.: -0.007 min
Response: 68321
Conc: 6.10 ng/ml