

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090924\
 Data File : P0106345.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2024 00:30
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO090924AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 04:56:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 04:31:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.451	3.714	409.7E6	152.6E6	53.075	54.163
2) SA Decachlor...	10.192	8.742	276.0E6	118.8E6	53.073	54.839
Target Compounds						
26) L6 AR-1254-1	6.469	5.607	133.5E6	75227911	512.265	534.071
27) L6 AR-1254-2	6.687	5.754	198.5E6	66886746	509.543	524.981
28) L6 AR-1254-3	7.052	6.159	201.3E6	107.6E6	516.211	534.429
29) L6 AR-1254-4	7.337	6.387	151.7E6	66739317	522.190	536.311
30) L6 AR-1254-5	7.755	6.805	165.4E6	96046707	543.627	536.242



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090924\
Data File : PO106345.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2024 00:30
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO090924AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 04:56:15 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 04:31:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.451 min
Delta R.T.: 0.000 min
Response: 409748377
Conc: 53.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO090924AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.714 min
Delta R.T.: 0.000 min
Response: 152649733
Conc: 54.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.192 min
Delta R.T.: -0.002 min
Response: 275956148
Conc: 53.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.742 min
Delta R.T.: 0.000 min
Response: 118802248
Conc: 54.84 ng/ml



#26 AR-1254-1


R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 133500374
Conc: 512.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO090924AR1254




#26 AR-1254-1

R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 75227911
Conc: 534.07 ng/ml




#27 AR-1254-2

R.T.: 6.687 min
Delta R.T.: 0.000 min
Response: 198542669
Conc: 509.54 ng/ml



#27 AR-1254-2


R.T.: 5.754 min
Delta R.T.: -0.001 min
Response: 66886746
Conc: 524.98 ng/ml



#28 AR-1254-3


R.T.: 7.052 min
Delta R.T.: 0.000 min
Response: 201270811
Conc: 516.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO090924AR1254




#28 AR-1254-3

R.T.: 6.159 min
Delta R.T.: -0.001 min
Response: 107595758
Conc: 534.43 ng/ml




#29 AR-1254-4

R.T.: 7.337 min
Delta R.T.: 0.000 min
Response: 151742998
Conc: 522.19 ng/ml



#29 AR-1254-4


R.T.: 6.387 min
Delta R.T.: 0.000 min
Response: 66739317
Conc: 536.31 ng/ml



#30 AR-1254-5

R.T.: 7.755 min
Delta R.T.: -0.002 min
Response: 165447635
Conc: 543.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO090924AR1254



#30 AR-1254-5

R.T.: 6.805 min
Delta R.T.: -0.001 min
Response: 96046707
Conc: 536.24 ng/ml