

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090924\
 Data File : P0106322.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2024 17:50
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 03:47:59 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 03:43:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.454	3.717	575.9E6	212.3E6	74.917	75.218
2)	SA Decachlor...	10.198	8.746	369.3E6	154.2E6	74.293	75.156
Target Compounds							
16)	L4 AR-1242-1	5.618	4.805	146.2E6	56109311	750.785	753.078
17)	L4 AR-1242-2	5.640	4.825	202.0E6	74572129	743.945	748.384
18)	L4 AR-1242-3	5.702	5.003	125.4E6	40294107	745.610	749.307
19)	L4 AR-1242-4	5.801	5.085	102.5E6	39116870	746.736	748.238
20)	L4 AR-1242-5	6.535	5.610	102.8E6	51557448	750.131	749.350



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090924\
Data File : PO106322.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2024 17:50
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 03:47:59 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 03:43:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.454 min
Delta R.T.: 0.000 min
Response: 575888582
Conc: 74.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: 0.000 min
Response: 212322738
Conc: 75.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.198 min
Delta R.T.: 0.000 min
Response: 369268522
Conc: 74.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.746 min
Delta R.T.: 0.000 min
Response: 154162451
Conc: 75.16 ng/ml



#16 AR-1242-1


R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 146236540
Conc: 750.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 56109311
Conc: 753.08 ng/ml




#17 AR-1242-2

R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 201977880
Conc: 743.94 ng/ml



#17 AR-1242-2


R.T.: 4.825 min
Delta R.T.: 0.000 min
Response: 74572129
Conc: 748.38 ng/ml



#18 AR-1242-3


R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 125372859
Conc: 745.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.003 min
Delta R.T.: 0.000 min
Response: 40294107
Conc: 749.31 ng/ml




#19 AR-1242-4

R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 102477682
Conc: 746.74 ng/ml



#19 AR-1242-4


R.T.: 5.085 min
Delta R.T.: 0.000 min
Response: 39116870
Conc: 748.24 ng/ml



#20 AR-1242-5

R.T.: 6.535 min
Delta R.T.: 0.000 min
Response: 102782444
Conc: 750.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 51557448
Conc: 749.35 ng/ml