

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0090924\
 Data File : P0106326.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2024 19:00
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 04:22:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 04:20:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.453	3.716	757.9E6	290.9E6	96.518	99.742
2) SA Decachlor...	10.197	8.745	495.2E6	207.5E6	93.586	94.026
Target Compounds						
21) L5 AR-1248-1	5.616	4.805	146.1E6	56436899	917.157	957.597
22) L5 AR-1248-2	5.889	5.042	200.4E6	74576878	913.674	941.806
23) L5 AR-1248-3	6.094	5.084	221.1E6	78365806	924.190	943.258
24) L5 AR-1248-4	6.496	5.257	245.4E6	94966635	934.938	957.709
25) L5 AR-1248-5	6.535	5.650	236.9E6	94802257	936.570	962.263



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO090924\
Data File : PO106326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2024 19:00
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 04:22:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 04:20:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: 0.000 min
Response: 757879533
Conc: 96.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.716 min
Delta R.T.: 0.000 min
Response: 290908513
Conc: 99.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.197 min
Delta R.T.: 0.000 min
Response: 495249516
Conc: 93.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.745 min
Delta R.T.: 0.000 min
Response: 207475768
Conc: 94.03 ng/ml



#21 AR-1248-1


R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 146113351
Conc: 917.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 56436899
Conc: 957.60 ng/ml




#22 AR-1248-2

R.T.: 5.889 min
Delta R.T.: 0.000 min
Response: 200351213
Conc: 913.67 ng/ml



#22 AR-1248-2


R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 74576878
Conc: 941.81 ng/ml



#23 AR-1248-3


R.T.: 6.094 min
Delta R.T.: 0.000 min
Response: 221131038
Conc: 924.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 78365806
Conc: 943.26 ng/ml




#24 AR-1248-4

R.T.: 6.496 min
Delta R.T.: 0.000 min
Response: 245431646
Conc: 934.94 ng/ml



#24 AR-1248-4


R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 94966635
Conc: 957.71 ng/ml



#25 AR-1248-5

R.T.: 6.535 min
Delta R.T.: 0.000 min
Response: 236852991
Conc: 936.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 94802257
Conc: 962.26 ng/ml