

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091019\
 Data File : P0060630.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Sep 2019 14:28
 Operator : SM/AJ
 Sample : PB122950BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB122950BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 02:56:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Sep 08 02:12:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.371	3.533	861776	716104	20.769	23.315
2) SA Decachlor...	10.029	8.419	880604	657119	16.694	17.751
Target Compounds						
9) L2 AR-1221-2	4.676	3.823	9020	11416	25.838	38.167 #
10) L2 AR-1221-3	4.676	0.000	9020	0	7.897	N.D. #
11) L3 AR-1232-1	4.676	0.000	9020	0	7.300	N.D. #
20) L4 AR-1242-5	6.476	0.000	13373	0	12.269	N.D. #
24) L5 AR-1248-4	6.476	0.000	13373	0	7.603	N.D. #
25) L5 AR-1248-5	6.476	0.000	13373	0	7.994	N.D. #
26) L6 AR-1254-1	6.320	0.000	5556	0	3.046	N.D. #
29) L6 AR-1254-4	7.243	6.202	8613	2797	3.587	1.565 #
30) L6 AR-1254-5	7.659	0.000	9690	0	3.562	N.D. #
32) L7 AR-1260-2	0.000	6.202	0	2797	N.D.	1.117 #
34) L7 AR-1260-4	0.000	6.783	0	259803	N.D.	127.915 #
36) L8 AR-1262-1	0.000	6.643	0	8990	N.D.	3.381 #
38) L8 AR-1262-3	8.581	0.000	2146	0	0.536	N.D. #
39) L8 AR-1262-4	0.000	7.456	0	17285	N.D.	5.040 #
41) L9 AR-1268-1	8.581	0.000	2146	0	0.311	N.D. #
42) L9 AR-1268-2	0.000	7.456	0	17285	N.D.	3.721 #
43) L9 AR-1268-3	8.911	0.000	32403	0	6.112	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091019\
Data File : PO060630.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Sep 2019 14:28
Operator : SM/AJ
Sample : PB122950BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB122950BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 02:56:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Sep 08 02:12:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.371 min
Delta R.T.: 0.002 min
Response: 861776
Conc: 20.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#1 Tetrachloro-m-xylene

R.T.: 3.533 min
Delta R.T.: -0.003 min
Response: 716104
Conc: 23.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.029 min
Delta R.T.: -0.008 min
Response: 880604
Conc: 16.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.419 min
Delta R.T.: -0.012 min
Response: 657119
Conc: 17.75 ng/ml



#9 AR-1221-2


R.T.: 4.676 min
Delta R.T.: 0.012 min
Response: 9020
Conc: 25.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#9 AR-1221-2

R.T.: 3.823 min
Delta R.T.: -0.007 min
Response: 11416
Conc: 38.17 ng/ml




#10 AR-1221-3

R.T.: 4.676 min
Delta R.T.: -0.064 min
Response: 9020
Conc: 7.90 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 3.904 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.676 min
Delta R.T.: -0.064 min
Response: 9020
Conc: 7.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.904 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.476 min
Delta R.T.: 0.026 min
Response: 13373
Conc: 12.27 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T. : 5.376 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.476 min
Delta R.T.: 0.065 min
Response: 13373
Conc: 7.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.032 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.476 min
Delta R.T.: 0.027 min
Response: 13373
Conc: 7.99 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 5.415 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.320 min
Delta R.T.: -0.065 min
Response: 5556
Conc: 3.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 5.377 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.243 min
Delta R.T.: -0.010 min
Response: 8613
Conc: 3.59 ng/ml



#29 AR-1254-4


R.T.: 6.202 min
Delta R.T.: 0.061 min
Response: 2797
Conc: 1.57 ng/ml



#30 AR-1254-5


R.T.: 7.659 min
Delta R.T.: -0.012 min
Response: 9690
Conc: 3.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.553 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.388 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.202 min
Delta R.T.: -0.028 min
Response: 2797
Conc: 1.12 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 7.971 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#34 AR-1260-4

R.T.: 6.783 min
Delta R.T.: -0.065 min
Response: 259803
Conc: 127.92 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.745 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.643 min
Delta R.T.: 0.051 min
Response: 8990
Conc: 3.38 ng/ml



#38 AR-1262-3


R.T.: 8.581 min
Delta R.T.: -0.008 min
Response: 2146
Conc: 0.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 7.367 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 8.676 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 7.456 min
Delta R.T.: 0.027 min
Response: 17285
Conc: 5.04 ng/ml



#41 AR-1268-1


R.T.: 8.581 min
Delta R.T.: -0.004 min
Response: 2146
Conc: 0.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.367 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.677 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.456 min
Delta R.T.: 0.026 min
Response: 17285
Conc: 3.72 ng/ml



#43 AR-1268-3

R.T.: 8.911 min
Delta R.T.: 0.013 min
Response: 32403
Conc: 6.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB122950BL



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.636 min
Response: 0
Conc: N.D.