

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0091020\  
Data File : P0071253.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 10 Sep 2020 9:41  
Operator : DD\AJ  
Sample : PB131491BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
PB131491BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 10 15:35:54 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0090920.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Sep 09 16:04:54 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1  | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-------|-------|---------|---------|--------|--------|
| -----                       |       |       |         |         |        |        |
| System Monitoring Compounds |       |       |         |         |        |        |
| 1) SA Tetrachlo...          | 4.351 | 3.531 | 1362400 | 743792  | 19.225 | 19.051 |
| 2) SA Decachlor...          | 9.962 | 8.710 | 1946843 | 1158195 | 20.877 | 21.026 |

Target Compounds

-----



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0091020\  
Data File : P0071253.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 10 Sep 2020 9:41  
Operator : DD\AJ  
Sample : PB131491BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
PB131491BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 10 15:35:54 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0090920.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Sep 09 16:04:54 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.351 min  
Delta R.T.: 0.007 min  
Response: 1362400  
Conc: 19.22 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
PB131491BL




#1 Tetrachloro-m-xylene

R.T.: 3.531 min  
Delta R.T.: 0.003 min  
Response: 743792  
Conc: 19.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.962 min  
Delta R.T.: 0.004 min  
Response: 1946843  
Conc: 20.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.710 min  
Delta R.T.: 0.002 min  
Response: 1158195  
Conc: 21.03 ng/ml