

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091021\
 Data File : P0081085.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Sep 2021 00:40
 Operator : DD\AJ
 Sample : M3684-15
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 01:12:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.926	4.053	1935873	586160	24.523	23.720
2)	SA Decachlor...	10.991	9.412	1576947	593228	26.559	25.412
Target Compounds							
8)	L2 AR-1221-1	5.180	0.000	123410	0	127.003	N.D. #
9)	L2 AR-1221-2	5.279	4.412	19969	10410	27.164	42.230 #
12)	L3 AR-1232-2	5.984f	0.000	49867	0	50.797	N.D. #
24)	L5 AR-1248-4	7.180f	0.000	9179	0	3.187	N.D. #
26)	L6 AR-1254-1	7.180	0.000	9179	0	2.840	N.D. #
28)	L6 AR-1254-3	7.805	0.000	37450	0	7.407	N.D. #
33)	L7 AR-1260-3	0.000	7.290	0	52726	N.D.	28.801 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091021\
Data File : PO081085.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Sep 2021 00:40
Operator : DD\AJ
Sample : M3684-15
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 01:12:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.926 min
Delta R.T.: -0.004 min
Response: 1935873
Conc: 24.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: -0.003 min
Response: 586160
Conc: 23.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.991 min
Delta R.T.: -0.008 min
Response: 1576947
Conc: 26.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.412 min
Delta R.T.: -0.008 min
Response: 593228
Conc: 25.41 ng/ml



#8 AR-1221-1


R.T.: 5.180 min
Delta R.T.: -0.003 min
Response: 123410
Conc: 127.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.317 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.279 min
Delta R.T.: -0.003 min
Response: 19969
Conc: 27.16 ng/ml



#9 AR-1221-2


R.T.: 4.412 min
Delta R.T.: -0.006 min
Response: 10410
Conc: 42.23 ng/ml



#12 AR-1232-2


R.T.: 5.984 min
Delta R.T.: 0.022 min
Response: 49867
Conc: 50.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.348 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.180 min
Delta R.T.: -0.027 min
Response: 9179
Conc: 3.19 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.825 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 7.180 min
Delta R.T.: 0.003 min
Response: 9179
Conc: 2.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.209 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.805 min
Delta R.T.: 0.013 min
Response: 37450
Conc: 7.41 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.795 min
Response: 0
Conc: N.D.



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 8.592 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#33 AR-1260-3

R.T.: 7.290 min
Delta R.T.: 0.005 min
Response: 52726
Conc: 28.80 ng/ml