

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091021\
 Data File : P0081135.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Sep 2021 17:40
 Operator : DD\AJ
 Sample : M3696-02 5X
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 H1H012E-END-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 01:40:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.928	4.054	244808	93409	3.101	3.780
2)	SA Decachlor...	10.993	9.416	206799	87771	3.483	3.760
Target Compounds							
26)	L6 AR-1254-1	7.173	6.205	175848	110545	54.412	69.882 #
27)	L6 AR-1254-2	7.404	6.365	292929	103733	58.857	71.749
28)	L6 AR-1254-3	7.788	6.790	301412	143662	59.610	64.212
29)	L6 AR-1254-4	8.085	7.031	224002	96448	59.941	68.133
30)	L6 AR-1254-5	8.515	7.463	230869	139826	55.812	66.270



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091021\
Data File : P0081135.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Sep 2021 17:40
Operator : DD\AJ
Sample : M3696-02 5X
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HHH012E-END-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 01:40:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.928 min
Delta R.T.: -0.002 min
Response: 244808
Conc: 3.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIH012E-END-1-2




#1 Tetrachloro-m-xylene

R.T.: 4.054 min
Delta R.T.: -0.001 min
Response: 93409
Conc: 3.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.993 min
Delta R.T.: -0.006 min
Response: 206799
Conc: 3.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.416 min
Delta R.T.: -0.004 min
Response: 87771
Conc: 3.76 ng/ml



#26 AR-1254-1


R.T.: 7.173 min
Delta R.T.: -0.004 min
Response: 175848
Conc: 54.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIH012E-END-1-2




#26 AR-1254-1

R.T.: 6.205 min
Delta R.T.: -0.004 min
Response: 110545
Conc: 69.88 ng/ml




#27 AR-1254-2

R.T.: 7.404 min
Delta R.T.: -0.004 min
Response: 292929
Conc: 58.86 ng/ml



#27 AR-1254-2


R.T.: 6.365 min
Delta R.T.: -0.004 min
Response: 103733
Conc: 71.75 ng/ml



#28 AR-1254-3


R.T.: 7.788 min
Delta R.T.: -0.004 min
Response: 301412
Conc: 59.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
HHH012E-END-1-2




#28 AR-1254-3

R.T.: 6.790 min
Delta R.T.: -0.004 min
Response: 143662
Conc: 64.21 ng/ml




#29 AR-1254-4

R.T.: 8.085 min
Delta R.T.: -0.004 min
Response: 224002
Conc: 59.94 ng/ml



#29 AR-1254-4


R.T.: 7.031 min
Delta R.T.: -0.004 min
Response: 96448
Conc: 68.13 ng/ml



#30 AR-1254-5

R.T.: 8.515 min
Delta R.T.: -0.004 min
Response: 230869
Conc: 55.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIH012E-END-1-2



#30 AR-1254-5

R.T.: 7.463 min
Delta R.T.: -0.005 min
Response: 139826
Conc: 66.27 ng/ml