

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091024\
 Data File : P0106361.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2024 14:50
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 02:18:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.453	3.717	365.9E6	137.9E6	48.782	51.633
2) SA Decachlor...	10.196	8.743	227.1E6	96633458	45.938	48.311
Target Compounds						
3) L1 AR-1016-1	5.616	4.804	113.0E6	47672655	442.350	500.486
4) L1 AR-1016-2	5.639	4.824	158.0E6	63276936	443.986	495.081
5) L1 AR-1016-3	5.702	5.001	96855233	33758572	444.396	489.378
6) L1 AR-1016-4	5.800	5.042	78418780	25862285	442.141	486.980
7) L1 AR-1016-5	6.094	5.256	76051374	37778818	440.648	528.851
31) L7 AR-1260-1	7.217	6.290	124.3E6	61647653	440.509	476.690
32) L7 AR-1260-2	7.473	6.477	147.3E6	74628190	441.586	470.220
33) L7 AR-1260-3	7.832	6.632	95354822	70021522	422.462	453.961
34) L7 AR-1260-4	8.057	7.104	116.0E6	47827607	419.097	449.861
35) L7 AR-1260-5	8.377	7.343	216.4E6	118.0E6	438.482	458.599



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091024\
Data File : PO106361.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2024 14:50
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 02:18:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: 0.000 min
Response: 365933690
Conc: 48.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: 0.000 min
Response: 137922807
Conc: 51.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.196 min
Delta R.T.: -0.002 min
Response: 227062460
Conc: 45.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.743 min
Delta R.T.: -0.003 min
Response: 96633458
Conc: 48.31 ng/ml



#3 AR-1016-1


R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 112984120
Conc: 442.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.804 min
Delta R.T.: -0.002 min
Response: 47672655
Conc: 500.49 ng/ml




#4 AR-1016-2

R.T.: 5.639 min
Delta R.T.: 0.000 min
Response: 158037009
Conc: 443.99 ng/ml



#4 AR-1016-2


R.T.: 4.824 min
Delta R.T.: -0.002 min
Response: 63276936
Conc: 495.08 ng/ml



#5 AR-1016-3


R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 96855233
Conc: 444.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.001 min
Delta R.T.: -0.002 min
Response: 33758572
Conc: 489.38 ng/ml




#6 AR-1016-4

R.T.: 5.800 min
Delta R.T.: 0.000 min
Response: 78418780
Conc: 442.14 ng/ml



#6 AR-1016-4


R.T.: 5.042 min
Delta R.T.: -0.001 min
Response: 25862285
Conc: 486.98 ng/ml



#7 AR-1016-5


R.T.: 6.094 min
Delta R.T.: -0.001 min
Response: 76051374
Conc: 440.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.256 min
Delta R.T.: -0.001 min
Response: 37778818
Conc: 528.85 ng/ml




#31 AR-1260-1

R.T.: 7.217 min
Delta R.T.: -0.002 min
Response: 124283171
Conc: 440.51 ng/ml



#31 AR-1260-1


R.T.: 6.290 min
Delta R.T.: -0.003 min
Response: 61647653
Conc: 476.69 ng/ml



#32 AR-1260-2


R.T.: 7.473 min
Delta R.T.: -0.002 min
Response: 147320627
Conc: 441.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.477 min
Delta R.T.: -0.003 min
Response: 74628190
Conc: 470.22 ng/ml




#33 AR-1260-3

R.T.: 7.832 min
Delta R.T.: -0.002 min
Response: 95354822
Conc: 422.46 ng/ml



#33 AR-1260-3


R.T.: 6.632 min
Delta R.T.: -0.003 min
Response: 70021522
Conc: 453.96 ng/ml



#34 AR-1260-4


R.T.: 8.057 min
Delta R.T.: -0.002 min
Response: 115953690
Conc: 419.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.104 min
Delta R.T.: -0.003 min
Response: 47827607
Conc: 449.86 ng/ml



#35 AR-1260-5

R.T.: 8.377 min
Delta R.T.: -0.002 min
Response: 216446968
Conc: 438.48 ng/ml



#35 AR-1260-5

R.T.: 7.343 min
Delta R.T.: -0.003 min
Response: 117961852
Conc: 458.60 ng/ml