

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091024\
 Data File : P0106363.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2024 15:25
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 02:19:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.453	3.717	385.8E6	147.0E6	51.433	55.014
2)	SA Decachlor...	10.196	8.743	246.6E6	106.8E6	49.882	53.396
Target Compounds							
21)	L5 AR-1248-1	5.616	4.804	74263109	29289978	470.244	506.283
22)	L5 AR-1248-2	5.889	5.041	98769038	38810121	455.307	496.523
23)	L5 AR-1248-3	6.094	5.083	106.6E6	40786831	446.370	496.263
24)	L5 AR-1248-4	6.496	5.256	122.7E6	48921698	467.280	498.264
25)	L5 AR-1248-5	6.534	5.649	119.3E6	47834838	475.633	493.886



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091024\
Data File : PO106363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2024 15:25
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 02:19:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: -0.001 min
Response: 385819228
Conc: 51.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.717 min
Delta R.T.: 0.000 min
Response: 146953651
Conc: 55.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.196 min
Delta R.T.: -0.001 min
Response: 246556882
Conc: 49.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.743 min
Delta R.T.: -0.004 min
Response: 106804310
Conc: 53.40 ng/ml



#21 AR-1248-1


R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 74263109
Conc: 470.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 29289978
Conc: 506.28 ng/ml




#22 AR-1248-2

R.T.: 5.889 min
Delta R.T.: 0.000 min
Response: 98769038
Conc: 455.31 ng/ml



#22 AR-1248-2


R.T.: 5.041 min
Delta R.T.: 0.000 min
Response: 38810121
Conc: 496.52 ng/ml



#23 AR-1248-3


R.T.: 6.094 min
Delta R.T.: 0.000 min
Response: 106589416
Conc: 446.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.083 min
Delta R.T.: 0.000 min
Response: 40786831
Conc: 496.26 ng/ml




#24 AR-1248-4

R.T.: 6.496 min
Delta R.T.: 0.000 min
Response: 122703615
Conc: 467.28 ng/ml



#24 AR-1248-4


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 48921698
Conc: 498.26 ng/ml



#25 AR-1248-5

R.T.: 6.534 min
Delta R.T.: 0.000 min
Response: 119277269
Conc: 475.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.649 min
Delta R.T.: 0.000 min
Response: 47834838
Conc: 493.89 ng/ml