

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091024\
 Data File : P0106364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2024 15:42
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 02:20:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.453	3.716	383.2E6	144.9E6	51.082	54.261
2)	SA Decachlor...	10.195	8.742	245.0E6	105.8E6	49.558	52.910
Target Compounds							
26)	L6 AR-1254-1	6.470	5.608	120.9E6	71946697	464.007	510.776
27)	L6 AR-1254-2	6.688	5.755	180.2E6	63467740	462.452	498.146
28)	L6 AR-1254-3	7.053	6.160	173.2E6	97584062	444.160	484.701
29)	L6 AR-1254-4	7.338	6.387	130.8E6	59269287	450.286	476.282
30)	L6 AR-1254-5	7.757	6.806	140.5E6	83824567	461.586	468.004



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091024\
Data File : PO106364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2024 15:42
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 02:20:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.453 min
Delta R.T.: -0.001 min
Response: 383184992
Conc: 51.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.716 min
Delta R.T.: -0.001 min
Response: 144942894
Conc: 54.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.195 min
Delta R.T.: -0.002 min
Response: 244955877
Conc: 49.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.742 min
Delta R.T.: -0.004 min
Response: 105833252
Conc: 52.91 ng/ml



#26 AR-1254-1


R.T.: 6.470 min
Delta R.T.: 0.001 min
Response: 120923831
Conc: 464.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.608 min
Delta R.T.: 0.000 min
Response: 71946697
Conc: 510.78 ng/ml




#27 AR-1254-2

R.T.: 6.688 min
Delta R.T.: 0.001 min
Response: 180193608
Conc: 462.45 ng/ml



#27 AR-1254-2


R.T.: 5.755 min
Delta R.T.: 0.000 min
Response: 63467740
Conc: 498.15 ng/ml



#28 AR-1254-3


R.T.: 7.053 min
Delta R.T.: 0.000 min
Response: 173177790
Conc: 444.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.160 min
Delta R.T.: 0.000 min
Response: 97584062
Conc: 484.70 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.001 min
Response: 130848447
Conc: 450.29 ng/ml



#29 AR-1254-4


R.T.: 6.387 min
Delta R.T.: 0.000 min
Response: 59269287
Conc: 476.28 ng/ml



#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: 0.000 min
Response: 140479312
Conc: 461.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.806 min
Delta R.T.: 0.000 min
Response: 83824567
Conc: 468.00 ng/ml