

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091025\
 Data File : P0113640.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2025 05:46
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/10/2025
 Supervised By :mohammad ahmed 09/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 06:32:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0081825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 09 15:11:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.659	3.649	476.4E6	277.1E6	58.195	54.761
2) SA Decachlor...	8.669	8.613	368.5E6	104.2E6	53.165	54.985
Target Compounds						
3) L1 AR-1016-1	4.741	4.721	148.6E6	87306114	556.771	511.838
4) L1 AR-1016-2	4.759	4.739	233.2E6	134.1E6	580.860	520.313
5) L1 AR-1016-3	4.817	4.914	146.3E6	73236029	573.870	524.559
6) L1 AR-1016-4	4.936	4.956	120.6E6	59589911	552.877	510.629
7) L1 AR-1016-5	5.192	5.167	125.0E6	78909017	564.154	553.631
31) L7 AR-1260-1	6.227	6.194	250.9E6	122.9E6	562.496	481.867
32) L7 AR-1260-2	6.416	6.383	393.8E6	161.2E6	573.505	466.220
33) L7 AR-1260-3	6.782	6.533	354.7E6	130.9E6	574.446	440.137
34) L7 AR-1260-4	7.041	7.001	255.4E6	93002442	570.304	448.746m
35) L7 AR-1260-5	7.284	7.244	677.2E6	202.3E6	559.358	431.919

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091025\
Data File : PO113640.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2025 05:46
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/10/2025
Supervised By :mohammad ahmed 09/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 06:32:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO081825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 09 15:11:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

