

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0091118.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 12 08:13:07 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0048899.D 1000 =P0048897.D 750 =P0048898.D
 250 =P0048900.D 50 =P0048901.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.603	1.646	1.649	1.715	1.579	1.638 E5	3.18
2) SA	Decachlorobiphenyl	2.229	2.003	2.193	2.326	2.388	2.228 E5	6.63
3) L1	AR-1016-1	4.347	4.375	4.546	4.683	4.714	4.533 E3	3.74
4) L1	AR-1016-2	4.529	4.597	4.720	5.022	4.855	4.745 E3	4.18
5) L1	AR-1016-3	6.876	6.910	7.149	7.542	7.667	7.229 E3	5.00
6) L1	AR-1016-4	6.066	6.059	6.359	6.622	6.330	6.287 E3	3.73
7) L1	AR-1016-5	5.289	5.306	5.516	5.768	4.698	5.316 E3	7.45
8) L2	AR-1221-1	1.788					1.788 E3	0.00
9) L2	AR-1221-2	1.369					1.369 E3	0.00
10) L2	AR-1221-3	4.523					4.523 E3	0.00
11) L3	AR-1232-1	3.621					3.621 E2	0.00
12) L3	AR-1232-2	1.920					1.920 E3	0.00
13) L3	AR-1232-3	5.393					5.393 E2	0.00
14) L3	AR-1232-4	4.814					4.814 E2	0.00
15) L3	AR-1232-5	3.089					3.089 E2	0.00
16) L4	AR-1242-1	1.671					1.671 E3	0.00
17) L4	AR-1242-2	3.254					3.254 E3	0.00
18) L4	AR-1242-3	1.569					1.569 E3	0.00
19) L4	AR-1242-4	1.202					1.202 E3	0.00
20) L4	AR-1242-5	4.343					4.343 E3	0.00
21) L5	AR-1248-1	5.563					5.563 E3	0.00
22) L5	AR-1248-2	6.085					6.085 E3	0.00
23) L5	AR-1248-3	8.026					8.026 E3	0.00
24) L5	AR-1248-4	7.668					7.668 E3	0.00
25) L5	AR-1248-5	5.241					5.241 E3	0.00
26) L6	AR-1254-1	1.558					1.558 E4	0.00
27) L6	AR-1254-2	1.015					1.015 E4	0.00
28) L6	AR-1254-3	1.792					1.792 E4	0.00
29) L6	AR-1254-4	1.383					1.383 E4	0.00
30) L6	AR-1254-5	1.009					1.009 E4	0.00
31) L7	AR-1260-1	1.203	1.151	1.218	1.325	1.354	1.250 E4	6.86
32) L7	AR-1260-2	1.356	1.276	1.351	1.460	1.494	1.387 E4	6.36
33) L7	AR-1260-3	1.484	1.386	1.473	1.576	1.586	1.501 E4	5.50
34) L7	AR-1260-4	1.002	0.926	0.987	1.051	1.037	1.001 E4	4.88
35) L7	AR-1260-5	2.038	1.925	2.038	2.099	2.071	2.034 E4	3.25
36) L8	AR-1262-1	6.578					6.578 E3	0.00
37) L8	AR-1262-2	5.137					5.137 E3	0.00
38) L8	AR-1262-3	5.085					5.085 E3	0.00
39) L8	AR-1262-4	1.081					1.081 E4	0.00
40) L8	AR-1262-5	8.175					8.175 E3	0.00
41) L9	AR-1268-1	5.114					5.114 E3	0.00
42) L9	AR-1268-2	6.118					6.118 E3	0.00
43) L9	AR-1268-3	2.621					2.621 E4	0.00
44) L9	AR-1268-4	2.157					2.157 E4	0.00
45) L9	AR-1268-5	7.394					7.394 E4	0.00

Signal #2 Calibration Files

500 =P0048899.D 1000 =P0048897.D 750 =P0048898.D
 250 =P0048900.D 50 =P0048901.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.762	1.861	1.729	1.892	1.878	1.824 E5	4.05
2) SA	Decachlorobiphenyl	1.771	1.647	1.690	1.933	2.237	1.856 E5	12.89
3) L1	AR-1016-1	4.884	5.005	4.750	5.389	5.356	5.077 E3	5.61
4) L1	AR-1016-2	7.372	7.636	7.200	8.095	8.407	7.742 E3	6.48
5) L1	AR-1016-3	0.988	1.035	0.968	1.085	1.089	1.033 E4	5.34

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0091118.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 12 08:13:07 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0048899.D 1000 =P0048897.D 750 =P0048898.D
 250 =P0048900.D 50 =P0048901.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	7.106	7.267	6.923	7.794	8.082	7.435	E3	6.54
7)	L1 AR-1016-5	6.304	6.632	6.305	7.575	7.872	6.937	E3	10.63
8)	L2 AR-1221-1	2.031					2.031	E3	0.00
9)	L2 AR-1221-2	1.546					1.546	E3	0.00
10)	L2 AR-1221-3	4.766					4.766	E3	0.00
11)	L3 AR-1232-1	1.982					1.982	E3	0.00
12)	L3 AR-1232-2	6.771					6.771	E2	0.00
13)	L3 AR-1232-3	1.876					1.876	E3	0.00
14)	L3 AR-1232-4	2.097					2.097	E3	0.00
15)	L3 AR-1232-5	2.547					2.547	E2	0.00
16)	L4 AR-1242-1	1.888					1.888	E3	0.00
17)	L4 AR-1242-2	4.159					4.159	E3	0.00
18)	L4 AR-1242-3	1.464					1.464	E3	0.00
19)	L4 AR-1242-4	1.259					1.259	E3	0.00
20)	L4 AR-1242-5	3.385					3.385	E3	0.00
21)	L5 AR-1248-1	6.707					6.707	E3	0.00
22)	L5 AR-1248-2	1.271					1.271	E4	0.00
23)	L5 AR-1248-3	8.977					8.977	E3	0.00
24)	L5 AR-1248-4	8.116					8.116	E3	0.00
25)	L5 AR-1248-5	6.272					6.272	E3	0.00
26)	L6 AR-1254-1	1.326					1.326	E4	0.00
27)	L6 AR-1254-2	1.167					1.167	E4	0.00
28)	L6 AR-1254-3	1.477					1.477	E4	0.00
29)	L6 AR-1254-4	1.118					1.118	E4	0.00
30)	L6 AR-1254-5	2.093					2.093	E4	0.00
31)	L7 AR-1260-1	1.304	1.272	1.251	1.422	1.533	1.357	E4	8.76
32)	L7 AR-1260-2	1.581	1.525	1.501	1.735	2.139	1.696	E4	15.57
33)	L7 AR-1260-3	1.537	1.477	1.464	1.676	1.882	1.607	E4	10.90
34)	L7 AR-1260-4	1.271	1.199	1.217	1.383	1.582	1.330	E4	11.87
35)	L7 AR-1260-5	2.943	2.786	2.815	3.187	3.641	3.074	E4	11.52
36)	L8 AR-1262-1	7.492					7.492	E3	0.00
37)	L8 AR-1262-2	7.539					7.539	E3	0.00
38)	L8 AR-1262-3	8.217					8.217	E3	0.00
39)	L8 AR-1262-4	1.236					1.236	E4	0.00
40)	L8 AR-1262-5	1.016					1.016	E4	0.00
41)	L9 AR-1268-1	7.587					7.587	E3	0.00
42)	L9 AR-1268-2	3.326					3.326	E4	0.00
43)	L9 AR-1268-3	2.730					2.730	E4	0.00
44)	L9 AR-1268-4	7.746					7.746	E3	0.00
45)	L9 AR-1268-5	7.091					7.091	E4	0.00

(#) = Out of Range