

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091120\
 Data File : P0071299.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Sep 2020 17:29
 Operator : DD\AJ
 Sample : L3976-03
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BSY-WC-WATER-DRUM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 18:10:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 09 16:04:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.531	1401570	885166	19.778	22.673
2) SA Decachlor...	9.960	8.711	1101591	764869	11.813	13.885

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091120\
Data File : P0071299.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Sep 2020 17:29
Operator : DD\AJ
Sample : L3976-03
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BSY-WC-WATER-DRUM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 18:10:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090920.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 09 16:04:54 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: 0.002 min
Response: 1401570
Conc: 19.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
BSY-WC-WATER-DRUM




#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: 0.003 min
Response: 885166
Conc: 22.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.960 min
Delta R.T.: 0.001 min
Response: 1101591
Conc: 11.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.711 min
Delta R.T.: 0.002 min
Response: 764869
Conc: 13.89 ng/ml