

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0091123\
 Data File : P0097888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2023 15:39
 Operator : YP/AJ
 Sample : 04367-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 PILE-3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/12/2023
 Supervised By :Ankita Jodhani 09/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 17:12:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.441	3.622	65724624	17595168	16.497	18.373
2) SA Decachlor...	10.287	8.655	66887788	30993539	35.799	36.276
Target Compounds						
26) L6 AR-1254-1	6.482	5.518	50431814	16525577	411.360	303.986m#
27) L6 AR-1254-2	6.703	5.665	42735619	7748490	238.235	160.909m#
28) L6 AR-1254-3	7.073	6.071	77749759	22719304	439.046	307.510 #
29) L6 AR-1254-4	7.359	6.300	17435029	6293742	161.978	158.804
30) L6 AR-1254-5	7.791	6.720	47816797	44741579	397.700m	726.438 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091123\
Data File : PO097888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2023 15:39
Operator : YP/AJ
Sample : 04367-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PILE-3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/12/2023
Supervised By :Ankita Jodhani 09/12/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 17:12:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

