

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091123\
 Data File : P0097928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 05:19
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 05:42:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.441	3.622	183.8E6	51306877	46.141	53.574
2) SA Decachlor...	10.287	8.654	96640166	45190437	51.722	52.893
Target Compounds						
21) L5 AR-1248-1	5.619	4.711	32617700	11203617	480.707	545.387
22) L5 AR-1248-2	5.895	4.949	49272759	16200219	455.625	531.718
23) L5 AR-1248-3	6.101	4.991	51594380	17038078	459.691	534.997
24) L5 AR-1248-4	6.508	5.163	50683754	19969466	490.271	533.622
25) L5 AR-1248-5	6.547	5.557	50379598	18339717	451.945	550.374



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091123\
Data File : PO097928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 05:19
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 05:42:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.441 min
Delta R.T.: -0.002 min
Response: 183827434
Conc: 46.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.001 min
Response: 51306877
Conc: 53.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.287 min
Delta R.T.: -0.008 min
Response: 96640166
Conc: 51.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.654 min
Delta R.T.: -0.009 min
Response: 45190437
Conc: 52.89 ng/ml



#21 AR-1248-1


R.T.: 5.619 min
Delta R.T.: -0.003 min
Response: 32617700
Conc: 480.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.711 min
Delta R.T.: -0.003 min
Response: 11203617
Conc: 545.39 ng/ml




#22 AR-1248-2

R.T.: 5.895 min
Delta R.T.: -0.003 min
Response: 49272759
Conc: 455.63 ng/ml



#22 AR-1248-2


R.T.: 4.949 min
Delta R.T.: -0.003 min
Response: 16200219
Conc: 531.72 ng/ml



#23 AR-1248-3


R.T.: 6.101 min
Delta R.T.: -0.002 min
Response: 51594380
Conc: 459.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.991 min
Delta R.T.: -0.003 min
Response: 17038078
Conc: 535.00 ng/ml




#24 AR-1248-4

R.T.: 6.508 min
Delta R.T.: -0.004 min
Response: 50683754
Conc: 490.27 ng/ml



#24 AR-1248-4


R.T.: 5.163 min
Delta R.T.: -0.003 min
Response: 19969466
Conc: 533.62 ng/ml



#25 AR-1248-5

R.T.: 6.547 min
Delta R.T.: -0.004 min
Response: 50379598
Conc: 451.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.557 min
Delta R.T.: -0.004 min
Response: 18339717
Conc: 550.37 ng/ml