

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091219\
 Data File : P0060722.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Sep 2019 23:52
 Operator : SM/AJ
 Sample : K4864-04
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SPUD-4-091119

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 04:34:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Sep 08 02:12:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.367	3.532	667329	558442	16.083	18.182
2)	SA Decachlor...	10.027	8.414	815373	638421	15.457	17.246
Target Compounds							
9)	L2 AR-1221-2	0.000	3.821	0	9729	N.D.	32.528 #
38)	L8 AR-1262-3	0.000	7.416	0	5345	N.D.	2.820 #
39)	L8 AR-1262-4	0.000	7.416	0	5345	N.D.	1.559 #
41)	L9 AR-1268-1	0.000	7.416	0	5345	N.D.	1.052 #
42)	L9 AR-1268-2	0.000	7.416	0	5345	N.D.	1.151 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091219\
Data File : PO060722.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Sep 2019 23:52
Operator : SM/AJ
Sample : K4864-04
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SPUD-4-091119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 04:34:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Sep 08 02:12:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.367 min
Delta R.T.: -0.002 min
Response: 667329
Conc: 16.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
SPUD-4-091119




#1 Tetrachloro-m-xylene

R.T.: 3.532 min
Delta R.T.: -0.004 min
Response: 558442
Conc: 18.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.027 min
Delta R.T.: -0.010 min
Response: 815373
Conc: 15.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.414 min
Delta R.T.: -0.017 min
Response: 638421
Conc: 17.25 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.664 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SPUD-4-091119




#9 AR-1221-2

R.T.: 3.821 min
Delta R.T.: -0.009 min
Response: 9729
Conc: 32.53 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.589 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.416 min
Delta R.T.: 0.048 min
Response: 5345
Conc: 2.82 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 8.676 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SPUD-4-091119




#39 AR-1262-4

R.T.: 7.416 min
Delta R.T.: -0.014 min
Response: 5345
Conc: 1.56 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 8.585 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.416 min
Delta R.T.: 0.049 min
Response: 5345
Conc: 1.05 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.677 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SPUD-4-091119



#42 AR-1268-2

R.T.: 7.416 min
Delta R.T.: -0.014 min
Response: 5345
Conc: 1.15 ng/ml