

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091219\
Data File : P0060721.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Sep 2019 23:36
Operator : SM/AJ
Sample : K4864-03
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SPUD-3-091119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 04:34:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Sep 08 02:12:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.368	3.532	676952	560354	16.315	18.244
2) SA Decachlor...	10.026	8.414	781545	609965	14.816	16.478

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091219\
Data File : PO060721.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Sep 2019 23:36
Operator : SM/AJ
Sample : K4864-03
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SPUD-3-091119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 04:34:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Sep 08 02:12:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.368 min
Delta R.T.: -0.001 min
Response: 676952
Conc: 16.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
SPUD-3-091119




#1 Tetrachloro-m-xylene

R.T.: 3.532 min
Delta R.T.: -0.005 min
Response: 560354
Conc: 18.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.026 min
Delta R.T.: -0.012 min
Response: 781545
Conc: 14.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.414 min
Delta R.T.: -0.017 min
Response: 609965
Conc: 16.48 ng/ml