

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091219\
 Data File : P0060732.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Sep 2019 2:51
 Operator : SM/AJ
 Sample : K4854-06
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RI14-6-(0-6)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 04:39:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Sep 08 02:12:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.368	3.534	1043515	743235	25.150	24.198
2) SA Decachlor...	10.026	8.441	2864133	507573	54.296m	13.712m#
Target Compounds						
26) L6 AR-1254-1	6.381	5.369	170.8E6	155.6E6	93628.493	82029.977
27) L6 AR-1254-2	6.599	5.514	243.0E6	132.1E6	83836.409	77292.715
28) L6 AR-1254-3	6.964	5.909	252.6E6	208.2E6	79200.749	74241.218
29) L6 AR-1254-4	7.249	6.133	205.2E6	141.0E6	85447.325	78897.553
30) L6 AR-1254-5	7.666	6.544	219.5E6	201.0E6	80692.166	76609.360



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091219\
Data File : PO060732.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Sep 2019 2:51
Operator : SM/AJ
Sample : K4854-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
RI14-6-(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 04:39:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090719.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Sep 08 02:12:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.368 min
Delta R.T.: 0.000 min
Response: 1043515
Conc: 25.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI14-6-(0-6)




#1 Tetrachloro-m-xylene

R.T.: 3.534 min
Delta R.T.: -0.002 min
Response: 743235
Conc: 24.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.026 min
Delta R.T.: -0.011 min
Response: 2864133
Conc: 54.30 ng/ml m



#2 Decachlorobiphenyl


R.T.: 8.441 min
Delta R.T.: 0.010 min
Response: 507573
Conc: 13.71 ng/ml m



#26 AR-1254-1


R.T.: 6.381 min
Delta R.T.: -0.003 min
Response: 170752086
Conc: 93628.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
R114-6-(0-6)




#26 AR-1254-1

R.T.: 5.369 min
Delta R.T.: -0.007 min
Response: 155640170
Conc: 82029.98 ng/ml




#27 AR-1254-2

R.T.: 6.599 min
Delta R.T.: -0.004 min
Response: 242971432
Conc: 83836.41 ng/ml



#27 AR-1254-2


R.T.: 5.514 min
Delta R.T.: -0.007 min
Response: 132090695
Conc: 77292.72 ng/ml



#28 AR-1254-3


R.T.: 6.964 min
Delta R.T.: -0.004 min
Response: 252583395
Conc: 79200.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
R114-6-(0-6)




#28 AR-1254-3

R.T.: 5.909 min
Delta R.T.: -0.008 min
Response: 208215155
Conc: 74241.22 ng/ml




#29 AR-1254-4

R.T.: 7.249 min
Delta R.T.: -0.004 min
Response: 205187489
Conc: 85447.32 ng/ml



#29 AR-1254-4


R.T.: 6.133 min
Delta R.T.: -0.008 min
Response: 140953342
Conc: 78897.55 ng/ml



#30 AR-1254-5

R.T.: 7.666 min
Delta R.T.: -0.005 min
Response: 219505188
Conc: 80692.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
R114-6-(0-6)



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: -0.009 min
Response: 201023591
Conc: 76609.36 ng/ml