

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091223\
 Data File : P0097942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 09:54
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 14:44:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.440	3.621	187.5E6	52128964	47.060	54.432
2) SA Decachlor...	10.285	8.653	98218791	46725099	52.567	54.689
Target Compounds						
16) L4 AR-1242-1	5.617	4.709	46215942	15435431	496.188	550.025
17) L4 AR-1242-2	5.640	4.728	67205391	21381905	479.017	538.742
18) L4 AR-1242-3	5.702	4.905	42452110	11710016	490.517	554.137
19) L4 AR-1242-4	5.801	4.989	33157853	12348474	488.903	538.798
20) L4 AR-1242-5	6.544	5.515	33874067	14537530	481.388	558.035



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091223\
Data File : PO097942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 09:54
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 14:44:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.440 min
Delta R.T.: -0.003 min
Response: 187490873
Conc: 47.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.002 min
Response: 52128964
Conc: 54.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.285 min
Delta R.T.: -0.010 min
Response: 98218791
Conc: 52.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.653 min
Delta R.T.: -0.011 min
Response: 46725099
Conc: 54.69 ng/ml



#16 AR-1242-1


R.T.: 5.617 min
Delta R.T.: -0.004 min
Response: 46215942
Conc: 496.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.709 min
Delta R.T.: -0.005 min
Response: 15435431
Conc: 550.03 ng/ml




#17 AR-1242-2

R.T.: 5.640 min
Delta R.T.: -0.004 min
Response: 67205391
Conc: 479.02 ng/ml



#17 AR-1242-2


R.T.: 4.728 min
Delta R.T.: -0.005 min
Response: 21381905
Conc: 538.74 ng/ml



#18 AR-1242-3


R.T.: 5.702 min
Delta R.T.: -0.005 min
Response: 42452110
Conc: 490.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.905 min
Delta R.T.: -0.005 min
Response: 11710016
Conc: 554.14 ng/ml




#19 AR-1242-4

R.T.: 5.801 min
Delta R.T.: -0.004 min
Response: 33157853
Conc: 488.90 ng/ml



#19 AR-1242-4


R.T.: 4.989 min
Delta R.T.: -0.006 min
Response: 12348474
Conc: 538.80 ng/ml



#20 AR-1242-5

R.T.: 6.544 min
Delta R.T.: -0.005 min
Response: 33874067
Conc: 481.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.515 min
Delta R.T.: -0.006 min
Response: 14537530
Conc: 558.04 ng/ml