

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091223\
 Data File : P0097944.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 10:45
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 14:46:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.440	3.621	171.5E6	48039352	43.048	50.162
2) SA Decachlor...	10.286	8.653	91238653	43514480	48.831	50.931
Target Compounds						
26) L6 AR-1254-1	6.480	5.516	53949545	26802733	440.054	493.033
27) L6 AR-1254-2	6.701	5.664	78951720	23536026	440.126	488.761
28) L6 AR-1254-3	7.070	6.069	76677376	36594715	432.990	495.317
29) L6 AR-1254-4	7.359	6.299	49853702	20852216	463.160	526.144
30) L6 AR-1254-5	7.782	6.718	54182086	31215017	450.641	506.817



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091223\
Data File : PO097944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 10:45
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 14:46:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.440 min
Delta R.T.: -0.003 min
Response: 171506617
Conc: 43.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.002 min
Response: 48039352
Conc: 50.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.286 min
Delta R.T.: -0.009 min
Response: 91238653
Conc: 48.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.653 min
Delta R.T.: -0.011 min
Response: 43514480
Conc: 50.93 ng/ml



#26 AR-1254-1


R.T.: 6.480 min
Delta R.T.: -0.003 min
Response: 53949545
Conc: 440.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.516 min
Delta R.T.: -0.004 min
Response: 26802733
Conc: 493.03 ng/ml




#27 AR-1254-2

R.T.: 6.701 min
Delta R.T.: -0.003 min
Response: 78951720
Conc: 440.13 ng/ml



#27 AR-1254-2


R.T.: 5.664 min
Delta R.T.: -0.003 min
Response: 23536026
Conc: 488.76 ng/ml



#28 AR-1254-3


R.T.: 7.070 min
Delta R.T.: -0.003 min
Response: 76677376
Conc: 432.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.069 min
Delta R.T.: -0.004 min
Response: 36594715
Conc: 495.32 ng/ml




#29 AR-1254-4

R.T.: 7.359 min
Delta R.T.: -0.004 min
Response: 49853702
Conc: 463.16 ng/ml



#29 AR-1254-4


R.T.: 6.299 min
Delta R.T.: -0.004 min
Response: 20852216
Conc: 526.14 ng/ml



#30 AR-1254-5

R.T.: 7.782 min
Delta R.T.: -0.005 min
Response: 54182086
Conc: 450.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.718 min
Delta R.T.: -0.005 min
Response: 31215017
Conc: 506.82 ng/ml