

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091223\
 Data File : P0097950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2023 13:31
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 14:50:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 07 02:56:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.440	3.621	189.6E6	52161389	47.598	54.466
2)	SA Decachlor...	10.286	8.653	101.1E6	47202374	54.125	55.247
Target Compounds							
21)	L5 AR-1248-1	5.618	4.709	33999973	11359878	501.078	552.993
22)	L5 AR-1248-2	5.894	4.948	52019803	16393534	481.027	538.063
23)	L5 AR-1248-3	6.099	4.990	54719708	17321902	487.537	543.909
24)	L5 AR-1248-4	6.506	5.162	54363609	20380616	525.867	544.609
25)	L5 AR-1248-5	6.545	5.557	53955996	18741566	484.028	562.434



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091223\
Data File : PO097950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2023 13:31
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 14:50:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 07 02:56:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.440 min
Delta R.T.: -0.003 min
Response: 189634353
Conc: 47.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.002 min
Response: 52161389
Conc: 54.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.286 min
Delta R.T.: -0.009 min
Response: 101129410
Conc: 54.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.653 min
Delta R.T.: -0.010 min
Response: 47202374
Conc: 55.25 ng/ml



#21 AR-1248-1


R.T.: 5.618 min
Delta R.T.: -0.004 min
Response: 33999973
Conc: 501.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.709 min
Delta R.T.: -0.005 min
Response: 11359878
Conc: 552.99 ng/ml




#22 AR-1248-2

R.T.: 5.894 min
Delta R.T.: -0.004 min
Response: 52019803
Conc: 481.03 ng/ml



#22 AR-1248-2


R.T.: 4.948 min
Delta R.T.: -0.004 min
Response: 16393534
Conc: 538.06 ng/ml



#23 AR-1248-3


R.T.: 6.099 min
Delta R.T.: -0.004 min
Response: 54719708
Conc: 487.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.990 min
Delta R.T.: -0.004 min
Response: 17321902
Conc: 543.91 ng/ml




#24 AR-1248-4

R.T.: 6.506 min
Delta R.T.: -0.006 min
Response: 54363609
Conc: 525.87 ng/ml



#24 AR-1248-4


R.T.: 5.162 min
Delta R.T.: -0.005 min
Response: 20380616
Conc: 544.61 ng/ml



#25 AR-1248-5

R.T.: 6.545 min
Delta R.T.: -0.005 min
Response: 53955996
Conc: 484.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.557 min
Delta R.T.: -0.005 min
Response: 18741566
Conc: 562.43 ng/ml