

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091224\
 Data File : P0106495.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2024 22:44
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/16/2024
 Supervised By :Ankita Jodhani 09/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 05:17:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 09:57:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.714	398.1E6	135.1E6	53.066	50.589
2) SA Decachlor...	10.192	8.738	223.4E6	98401821	45.199	49.195
Target Compounds						
3) L1 AR-1016-1	5.615	4.801	125.2E6	47524049	490.088	498.926
4) L1 AR-1016-2	5.638	4.821	172.2E6	63326096	483.812	495.466
5) L1 AR-1016-3	5.699	4.998	105.8E6	34160012	485.525	495.197
6) L1 AR-1016-4	5.798	5.038	85404882	25468586	481.530	479.567
7) L1 AR-1016-5	6.092	5.252	85393847	36130271	494.779	505.774m
31) L7 AR-1260-1	7.216	6.287	134.2E6	65075816	475.566	503.198
32) L7 AR-1260-2	7.472	6.474	153.8E6	80482543	461.003	507.108
33) L7 AR-1260-3	7.830	6.628	100.4E6	75559535	444.606	489.865
34) L7 AR-1260-4	8.055	7.100	115.9E6	53267802	418.820	501.030
35) L7 AR-1260-5	8.374	7.340	212.3E6	126.7E6	430.058	492.697

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091224\
Data File : PO106495.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2024 22:44
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/16/2024
Supervised By :Ankita Jodhani 09/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 05:17:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 09:57:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

