Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091319\
Data File : PO060793.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Sep 2019 22:56
Operator : SM/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId : AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 02:43:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.365	3.530	2321886	1658743	63.110	59.406	
2)	SA Decachlor...	10.024	8.410	2542973	1895529	55.347	54.078	
	Target Compounds							
3)	L1 AR-1016-1	5.528	4.585	952963	702562	596.293	568.305	
4)	L1 AR-1016-2	5.550	4.603	1368974	977972	597.974	574.828	
5)	L1 AR-1016-3	5.612	4.773	831940	551342	579.514	568.704	
6)	L1 AR-1016-4	5.710	4.815	702115	459801	595.229	561.745	
7)	L1 AR-1016-5	6.003	5.022	739192	584084	589.677	557.669	
8)	L2 AR-1221-1	4.569	3.739	78309	65506	178.414	185.055	
9)	L2 AR-1221-2	4.661	3.822	119972	97093	351.248	345.478	
10)	L2 AR-1221-3	4.736	3.896	438489	344625	413.978	415.177	
11)	L3 AR-1232-1	4.736	3.896	438489	344625	375.814	378.150	
12)	L3 AR-1232-2	5.261	4.603	648343	977972	999.928	968.610	
13)	L3 AR-1232-3	5.550	4.773	1368974	551342	1022.116	982.433	
14)	L3 AR-1232-4	5.710	4.855	702115	564610	1021.790	1081.171	
15)	L3 AR-1232-5	5.799	5.022	646607	584084	1180.818	1032.075	
16)	L4 AR-1242-1	5.528	4.585	952963	702562	798.517	771.723	
17)	L4 AR-1242-2	5.550	4.603	1368974	977972	800.993	781.217	
18)	L4 AR-1242-3	5.612	4.773	831940	551342	768.265	771.586	
19)	L4 AR-1242-4	5.710	4.855	702115	564610	795.326	775.308	
20)	L4 AR-1242-5	6.444	5.365	135752	496229	121.133	501.290	#
21)	L5 AR-1248-1	5.528	4.585	952963	702562	1035.154	990.535	
22)	L5 AR-1248-2	5.799	4.815	646607	459801	483.534	475.888	
23)	L5 AR-1248-3	6.003	4.855	739192	564610	488.417	560.614	
24)	L5 AR-1248-4	6.404	5.022	165389	584084	90.780	486.362	#
25)	L5 AR-1248-5	6.444	5.403	135752	147683	77.591	116.823	#
26)	L6 AR-1254-1	6.378	5.365	616611	496229	320.044	240.778	
27)	L6 AR-1254-2	6.595	5.509	602342	483995	191.406	256.485	#
28)	L6 AR-1254-3	6.962	5.919	389070	887481	110.469	279.125	#
29)	L6 AR-1254-4	7.245	6.128	291831	145920	107.804	70.088	#
30)	L6 AR-1254-5	7.662	6.539	2060862	1562730	698.877	493.769	#
31)	L7 AR-1260-1	7.124	6.032	1447567	1134825	594.578	539.521	
32)	L7 AR-1260-2	7.379	6.217	1758295	1358061	581.572	526.483	
33)	L7 AR-1260-3	7.738	6.367	1370103	1287965	574.121	534.854	
34)	L7 AR-1260-4	7.963	6.832	1581129	1130120	565.155	528.034	
35)	L7 AR-1260-5	8.276	7.071	2966717	2532467	544.818	524.050	
36)	L8 AR-1262-1	7.738	6.577	1370103	1203616	387.878	402.965	
37)	L8 AR-1262-2	8.276	7.071	2966717	2532467	475.858	487.062	
38)	L8 AR-1262-3	8.594	7.351	1870821	625994	439.068	294.943	#
39)	L8 AR-1262-4	8.666	7.413	474651	1868956	145.580	489.322	#
40)	L8 AR-1262-5	9.303	7.902	812342	671798	399.503	401.366	
41)	L9 AR-1268-1	8.594	7.351	1870821	625994	253.483	106.788	#

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091319\
Data File : PO060793.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Sep 2019 22:56
Operator : SM/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument : ECD_O
ClientSampleId : AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 02:43:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.666	7.413	474651	1868956	70.266	350.650 #
43) L9	AR-1268-3	8.888	7.618	45117	46158	8.034	10.413 #
44) L9	AR-1268-4	9.303	7.902	812342	671798	352.162	351.569
45) L9	AR-1268-5	9.699	8.174	224461	192713	14.151	16.015

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091319\
Data File : PO060793.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Sep 2019 22:56
Operator : SM/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 02:43:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.365 min
Delta R.T.: -0.005 min
Response: 2321886
Conc: 63.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.001 min
Response: 1658743
Conc: 59.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.024 min
Delta R.T.: -0.004 min
Response: 2542973
Conc: 55.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.410 min
Delta R.T.: -0.003 min
Response: 1895529
Conc: 54.08 ng/ml



#3 AR-1016-1

R.T.: 5.528 min
Delta R.T.: -0.006 min
Response: 952963
Conc: 596.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#3 AR-1016-1

R.T.: 4.585 min
Delta R.T.: -0.001 min
Response: 702562
Conc: 568.31 ng/ml



#4 AR-1016-2

R.T.: 5.550 min
Delta R.T.: -0.006 min
Response: 1368974
Conc: 597.97 ng/ml



#4 AR-1016-2

R.T.: 4.603 min
Delta R.T.: -0.002 min
Response: 977972
Conc: 574.83 ng/ml



#5 AR-1016-3

R.T.: 5.612 min
Delta R.T.: -0.005 min
Response: 831940
Conc: 579.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#5 AR-1016-3

R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 551342
Conc: 568.70 ng/ml



#6 AR-1016-4

R.T.: 5.710 min
Delta R.T.: -0.005 min
Response: 702115
Conc: 595.23 ng/ml



#6 AR-1016-4

R.T.: 4.815 min
Delta R.T.: -0.002 min
Response: 459801
Conc: 561.74 ng/ml



#7 AR-1016-5

R.T.: 6.003 min
Delta R.T.: -0.006 min
Response: 739192
Conc: 589.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#7 AR-1016-5

R.T.: 5.022 min
Delta R.T.: -0.002 min
Response: 584084
Conc: 557.67 ng/ml



#8 AR-1221-1

R.T.: 4.569 min
Delta R.T.: -0.009 min
Response: 78309
Conc: 178.41 ng/ml



#8 AR-1221-1

R.T.: 3.739 min
Delta R.T.: -0.002 min
Response: 65506
Conc: 185.06 ng/ml



#9 AR-1221-2

R.T.: 4.661 min
Delta R.T.: -0.004 min
Response: 119972
Conc: 351.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#9 AR-1221-2

R.T.: 3.822 min
Delta R.T.: -0.001 min
Response: 97093
Conc: 345.48 ng/ml



#10 AR-1221-3

R.T.: 4.736 min
Delta R.T.: -0.005 min
Response: 438489
Conc: 413.98 ng/ml



#10 AR-1221-3

R.T.: 3.896 min
Delta R.T.: -0.001 min
Response: 344625
Conc: 415.18 ng/ml



#11 AR-1232-1

R.T.: 4.736 min
Delta R.T.: -0.005 min
Response: 438489
Conc: 375.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#11 AR-1232-1

R.T.: 3.896 min
Delta R.T.: -0.001 min
Response: 344625
Conc: 378.15 ng/ml



#12 AR-1232-2

R.T.: 5.261 min
Delta R.T.: -0.005 min
Response: 648343
Conc: 999.93 ng/ml



#12 AR-1232-2

R.T.: 4.603 min
Delta R.T.: -0.002 min
Response: 977972
Conc: 968.61 ng/ml



#13 AR-1232-3

R.T.: 5.550 min
Delta R.T.: -0.005 min
Response: 1368974
Conc: 1022.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#13 AR-1232-3

R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 551342
Conc: 982.43 ng/ml



#14 AR-1232-4

R.T.: 5.710 min
Delta R.T.: -0.005 min
Response: 702115
Conc: 1021.79 ng/ml



#14 AR-1232-4

R.T.: 4.855 min
Delta R.T.: -0.002 min
Response: 564610
Conc: 1081.17 ng/ml



#15 AR-1232-5

R.T.: 5.799 min
Delta R.T.: -0.005 min
Response: 646607
Conc: 1180.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#15 AR-1232-5

R.T.: 5.022 min
Delta R.T.: -0.002 min
Response: 584084
Conc: 1032.08 ng/ml



#16 AR-1242-1

R.T.: 5.528 min
Delta R.T.: -0.005 min
Response: 952963
Conc: 798.52 ng/ml



#16 AR-1242-1

R.T.: 4.585 min
Delta R.T.: -0.002 min
Response: 702562
Conc: 771.72 ng/ml



#17 AR-1242-2

R.T.: 5.550 min
Delta R.T.: -0.005 min
Response: 1368974
Conc: 800.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#17 AR-1242-2

R.T.: 4.603 min
Delta R.T.: -0.002 min
Response: 977972
Conc: 781.22 ng/ml



#18 AR-1242-3

R.T.: 5.612 min
Delta R.T.: -0.006 min
Response: 831940
Conc: 768.27 ng/ml



#18 AR-1242-3

R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 551342
Conc: 771.59 ng/ml



#19 AR-1242-4

R.T.: 5.710 min
Delta R.T.: -0.005 min
Response: 702115
Conc: 795.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#19 AR-1242-4

R.T.: 4.855 min
Delta R.T.: -0.003 min
Response: 564610
Conc: 775.31 ng/ml



#20 AR-1242-5

R.T.: 6.444 min
Delta R.T.: -0.004 min
Response: 135752
Conc: 121.13 ng/ml



#20 AR-1242-5

R.T.: 5.365 min
Delta R.T.: -0.001 min
Response: 496229
Conc: 501.29 ng/ml



#21 AR-1248-1

R.T.: 5.528 min
Delta R.T.: -0.005 min
Response: 952963
Conc: 1035.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#21 AR-1248-1

R.T.: 4.585 min
Delta R.T.: -0.001 min
Response: 702562
Conc: 990.53 ng/ml



#22 AR-1248-2

R.T.: 5.799 min
Delta R.T.: -0.004 min
Response: 646607
Conc: 483.53 ng/ml



#22 AR-1248-2

R.T.: 4.815 min
Delta R.T.: -0.002 min
Response: 459801
Conc: 475.89 ng/ml



#23 AR-1248-3

R.T.: 6.003 min
Delta R.T.: -0.004 min
Response: 739192
Conc: 488.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#23 AR-1248-3

R.T.: 4.855 min
Delta R.T.: -0.002 min
Response: 564610
Conc: 560.61 ng/ml



#24 AR-1248-4

R.T.: 6.404 min
Delta R.T.: -0.005 min
Response: 165389
Conc: 90.78 ng/ml



#24 AR-1248-4

R.T.: 5.022 min
Delta R.T.: -0.002 min
Response: 584084
Conc: 486.36 ng/ml



#25 AR-1248-5

R.T.: 6.444 min
Delta R.T.: -0.004 min
Response: 135752
Conc: 77.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#25 AR-1248-5

R.T.: 5.403 min
Delta R.T.: -0.004 min
Response: 147683
Conc: 116.82 ng/ml



#26 AR-1254-1

R.T.: 6.378 min
Delta R.T.: -0.004 min
Response: 616611
Conc: 320.04 ng/ml



#26 AR-1254-1

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 496229
Conc: 240.78 ng/ml



#27 AR-1254-2

R.T.: 6.595 min
Delta R.T.: -0.005 min
Response: 602342
Conc: 191.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#27 AR-1254-2

R.T.: 5.509 min
Delta R.T.: -0.001 min
Response: 483995
Conc: 256.49 ng/ml



#28 AR-1254-3

R.T.: 6.962 min
Delta R.T.: -0.003 min
Response: 389070
Conc: 110.47 ng/ml



#28 AR-1254-3

R.T.: 5.919 min
Delta R.T.: 0.015 min
Response: 887481
Conc: 279.12 ng/ml



#29 AR-1254-4

R.T.: 7.245 min
Delta R.T.: -0.005 min
Response: 291831
Conc: 107.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#29 AR-1254-4

R.T.: 6.128 min
Delta R.T.: -0.001 min
Response: 145920
Conc: 70.09 ng/ml



#30 AR-1254-5

R.T.: 7.662 min
Delta R.T.: -0.005 min
Response: 2060862
Conc: 698.88 ng/ml



#30 AR-1254-5

R.T.: 6.539 min
Delta R.T.: -0.002 min
Response: 1562730
Conc: 493.77 ng/ml



#31 AR-1260-1

R.T.: 7.124 min
Delta R.T.: -0.004 min
Response: 1447567
Conc: 594.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#31 AR-1260-1

R.T.: 6.032 min
Delta R.T.: -0.003 min
Response: 1134825
Conc: 539.52 ng/ml



#32 AR-1260-2

R.T.: 7.379 min
Delta R.T.: -0.005 min
Response: 1758295
Conc: 581.57 ng/ml



#32 AR-1260-2

R.T.: 6.217 min
Delta R.T.: -0.002 min
Response: 1358061
Conc: 526.48 ng/ml



#33 AR-1260-3

R.T.: 7.738 min
Delta R.T.: -0.004 min
Response: 1370103
Conc: 574.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#33 AR-1260-3

R.T.: 6.367 min
Delta R.T.: -0.002 min
Response: 1287965
Conc: 534.85 ng/ml



#34 AR-1260-4

R.T.: 7.963 min
Delta R.T.: -0.003 min
Response: 1581129
Conc: 565.15 ng/ml



#34 AR-1260-4

R.T.: 6.832 min
Delta R.T.: -0.003 min
Response: 1130120
Conc: 528.03 ng/ml



#35 AR-1260-5

R.T.: 8.276 min
Delta R.T.: -0.003 min
Response: 2966717
Conc: 544.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#35 AR-1260-5

R.T.: 7.071 min
Delta R.T.: -0.002 min
Response: 2532467
Conc: 524.05 ng/ml



#36 AR-1262-1

R.T.: 7.738 min
Delta R.T.: -0.008 min
Response: 1370103
Conc: 387.88 ng/ml



#36 AR-1262-1

R.T.: 6.577 min
Delta R.T.: -0.004 min
Response: 1203616
Conc: 402.96 ng/ml



#37 AR-1262-2

R.T.: 8.276 min
Delta R.T.: -0.008 min
Response: 2966717
Conc: 475.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#37 AR-1262-2

R.T.: 7.071 min
Delta R.T.: -0.004 min
Response: 2532467
Conc: 487.06 ng/ml



#38 AR-1262-3

R.T.: 8.594 min
Delta R.T.: 0.007 min
Response: 1870821
Conc: 439.07 ng/ml



#38 AR-1262-3

R.T.: 7.351 min
Delta R.T.: -0.005 min
Response: 625994
Conc: 294.94 ng/ml



#39 AR-1262-4

R.T.: 8.666 min
Delta R.T.: -0.008 min
Response: 474651
Conc: 145.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#39 AR-1262-4

R.T.: 7.413 min
Delta R.T.: -0.004 min
Response: 1868956
Conc: 489.32 ng/ml



#40 AR-1262-5

R.T.: 9.303 min
Delta R.T.: -0.008 min
Response: 812342
Conc: 399.50 ng/ml



#40 AR-1262-5

R.T.: 7.902 min
Delta R.T.: -0.004 min
Response: 671798
Conc: 401.37 ng/ml



#41 AR-1268-1

R.T.: 8.594 min
Delta R.T.: 0.014 min
Response: 1870821
Conc: 253.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#41 AR-1268-1

R.T.: 7.351 min
Delta R.T.: -0.003 min
Response: 625994
Conc: 106.79 ng/ml



#42 AR-1268-2

R.T.: 8.666 min
Delta R.T.: -0.008 min
Response: 474651
Conc: 70.27 ng/ml



#42 AR-1268-2

R.T.: 7.413 min
Delta R.T.: -0.003 min
Response: 1868956
Conc: 350.65 ng/ml



#43 AR-1268-3

R.T.: 8.888 min
Delta R.T.: -0.005 min
Response: 45117
Conc: 8.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#43 AR-1268-3

R.T.: 7.618 min
Delta R.T.: -0.004 min
Response: 46158
Conc: 10.41 ng/ml



#44 AR-1268-4

R.T.: 9.303 min
Delta R.T.: -0.003 min
Response: 812342
Conc: 352.16 ng/ml



#44 AR-1268-4

R.T.: 7.902 min
Delta R.T.: -0.002 min
Response: 671798
Conc: 351.57 ng/ml



#45 AR-1268-5

R.T.: 9.699 min
Delta R.T.: -0.004 min
Response: 224461
Conc: 14.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500



#45 AR-1268-5

R.T.: 8.174 min
Delta R.T.: -0.003 min
Response: 192713
Conc: 16.02 ng/ml