

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091321\
 Data File : P0081208.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Sep 2021 9:45
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 11:40:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.925	4.052	1588523	496844	20.123	20.105
2)	SA Decachlor...	10.987	9.411	1295625	503330	21.821	21.561
Target Compounds							
9)	L2 AR-1221-2	5.280	0.000	21474	0	29.211	N.D. #
20)	L4 AR-1242-5	7.218f	6.229f	4591	24253	2.865	33.488 #
25)	L5 AR-1248-5	7.218f	6.229f	4591	24253	1.669	26.751 #
26)	L6 AR-1254-1	7.178	6.229f	166804	24253	51.613	15.332 #
28)	L6 AR-1254-3	7.803	0.000	39264	0	7.765	N.D. #
32)	L7 AR-1260-2	0.000	7.101f	0	11567	N.D.	6.590 #
33)	L7 AR-1260-3	0.000	7.292	0	61745	N.D.	33.728 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091321\
Data File : PO081208.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Sep 2021 9:45
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 11:40:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.925 min
Delta R.T.: -0.005 min
Response: 1588523
Conc: 20.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.052 min
Delta R.T.: -0.003 min
Response: 496844
Conc: 20.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.987 min
Delta R.T.: -0.012 min
Response: 1295625
Conc: 21.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.411 min
Delta R.T.: -0.009 min
Response: 503330
Conc: 21.56 ng/ml



#9 AR-1221-2


R.T.: 5.280 min
Delta R.T.: -0.002 min
Response: 21474
Conc: 29.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.418 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 7.218 min
Delta R.T.: -0.029 min
Response: 4591
Conc: 2.87 ng/ml



#20 AR-1242-5


R.T.: 6.229 min
Delta R.T.: 0.020 min
Response: 24253
Conc: 33.49 ng/ml



#24 AR-1248-4


R.T.: 7.218 min
Delta R.T.: 0.011 min
Response: 4591
Conc: 1.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 5.843 min
Delta R.T.: 0.018 min
Response: -2359
Conc: N.D.




#25 AR-1248-5

R.T.: 7.218 min
Delta R.T.: -0.028 min
Response: 4591
Conc: 1.67 ng/ml



#25 AR-1248-5


R.T.: 6.229 min
Delta R.T.: -0.022 min
Response: 24253
Conc: 26.75 ng/ml



#26 AR-1254-1


R.T.: 7.178 min
Delta R.T.: 0.002 min
Response: 166804
Conc: 51.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 6.229 min
Delta R.T.: 0.019 min
Response: 24253
Conc: 15.33 ng/ml




#28 AR-1254-3

R.T.: 7.803 min
Delta R.T.: 0.012 min
Response: 39264
Conc: 7.77 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.795 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 8.226 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 7.101 min
Delta R.T.: -0.027 min
Response: 11567
Conc: 6.59 ng/ml



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.592 min
Response: 0
Conc: N.D.



#33 AR-1260-3

R.T.: 7.292 min
Delta R.T.: 0.007 min
Response: 61745
Conc: 33.73 ng/ml