

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091321\
 Data File : P0081211.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Sep 2021 10:37
 Operator : AJ\MA
 Sample : M3700-10
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RW-2-20210909

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 11:40:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.927	4.053	1378011	499423	17.456	20.210
2)	SA Decachlor...	10.987	9.412	812430	308188	13.683	13.202

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091321\
Data File : P0081211.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Sep 2021 10:37
Operator : AJ\MA
Sample : M3700-10
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
RW-2-20210909

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 11:40:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 1378011
Conc: 17.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
RW-2-20210909




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: -0.003 min
Response: 499423
Conc: 20.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.987 min
Delta R.T.: -0.012 min
Response: 812430
Conc: 13.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.412 min
Delta R.T.: -0.008 min
Response: 308188
Conc: 13.20 ng/ml