

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091521\
 Data File : P0081244.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Sep 2021 17:15
 Operator : AJ\MA
 Sample : M3730-02
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 WALLS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 16 05:43:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.927	4.054	1900406	606147	24.073	24.528
2)	SA Decachlor...	10.994	9.415	1529571	587683	25.761	25.175
Target Compounds							
9)	L2 AR-1221-2	5.282	4.410	23398	11151	31.829	45.236 #
31)	L7 AR-1260-1	7.960	0.000	15136	0	4.354	N.D. #
32)	L7 AR-1260-2	8.198f	0.000	128654	0	28.868	N.D. #
33)	L7 AR-1260-3	0.000	7.292	0	99636	N.D.	54.425 #
40)	L8 AR-1262-5	0.000	8.875	0	143621	N.D.	124.077 #
43)	L9 AR-1268-3	0.000	8.597f	0	280222	N.D.	94.862 #
44)	L9 AR-1268-4	0.000	8.875	0	143621	N.D.	109.147 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091521\
Data File : PO081244.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Sep 2021 17:15
Operator : AJ\MA
Sample : M3730-02
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
WALLS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 16 05:43:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 1900406
Conc: 24.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
WALLS




#1 Tetrachloro-m-xylene

R.T.: 4.054 min
Delta R.T.: -0.001 min
Response: 606147
Conc: 24.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.994 min
Delta R.T.: -0.005 min
Response: 1529571
Conc: 25.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.415 min
Delta R.T.: -0.005 min
Response: 587683
Conc: 25.17 ng/ml



#9 AR-1221-2


R.T.: 5.282 min
Delta R.T.: 0.000 min
Response: 23398
Conc: 31.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
WALLS




#9 AR-1221-2

R.T.: 4.410 min
Delta R.T.: -0.008 min
Response: 11151
Conc: 45.24 ng/ml




#31 AR-1260-1

R.T.: 7.960 min
Delta R.T.: 0.002 min
Response: 15136
Conc: 4.35 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.931 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.198 min
Delta R.T.: -0.028 min
Response: 128654
Conc: 28.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
WALLS




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.128 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.592 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.292 min
Delta R.T.: 0.007 min
Response: 99636
Conc: 54.43 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 10.236 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
WALLS




#40 AR-1262-5

R.T.: 8.875 min
Delta R.T.: 0.002 min
Response: 143621
Conc: 124.08 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 9.791 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.597 min
Delta R.T.: 0.016 min
Response: 280222
Conc: 94.86 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 10.238 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
WALLS



#44 AR-1268-4

R.T.: 8.875 min
Delta R.T.: 0.002 min
Response: 143621
Conc: 109.15 ng/ml