

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091522\
 Data File : P0089555.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2022 16:27
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 17:02:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 15 16:59:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.476	3.663	326.3E6	138.7E6	74.627	74.105
2) SA Decachlor...	10.360	8.776	202.8E6	87199964	73.952	71.513
Target Compounds						
16) L4 AR-1242-1	5.656	4.768	75405187	31020044	723.652	716.304
17) L4 AR-1242-2	5.679	4.788	111.4E6	45646324	741.876	723.343
18) L4 AR-1242-3	5.742	4.967	68250096	24364913	729.555	714.768
19) L4 AR-1242-4	5.842	5.052	55307601	25251312	734.472	707.569
20) L4 AR-1242-5	6.584	5.626	54728014	23649889	738.239	709.486

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091522\
Data File : PO089555.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2022 16:27
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 17:02:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091522.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 15 16:59:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

