

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091522\
 Data File : P0089547.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2022 14:13
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 15:58:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 15 15:56:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.476	3.663	397.0E6	169.3E6	98.024	96.097
2) SA Decachlor...	10.355	8.774	247.8E6	106.8E6	95.973	91.243
Target Compounds						
3) L1 AR-1016-1	5.654	4.768	112.5E6	44933361	952.158	914.680
4) L1 AR-1016-2	5.677	4.787	163.8E6	67872085	960.111	929.873
5) L1 AR-1016-3	5.741	4.966	100.5E6	35529637	946.562	912.203
6) L1 AR-1016-4	5.840	5.009	81241696	29433704	957.882	890.784
7) L1 AR-1016-5	6.137	5.226	78251058	38046245	945.779	907.886
31) L7 AR-1260-1	7.272	6.276	141.5E6	69356360	951.501	915.447
32) L7 AR-1260-2	7.530	6.466	162.2E6	80402414	952.291	920.359
33) L7 AR-1260-3	7.891	6.622	119.9E6	75910361	956.245	919.856
34) L7 AR-1260-4	8.120	7.101	135.8E6	60168947	962.819	916.461
35) L7 AR-1260-5	8.452	7.343	269.6E6	126.2E6	975.485	940.707



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091522\
Data File : P0089547.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2022 14:13
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 15:58:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091522.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 15 15:56:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.476 min
Delta R.T.: 0.000 min
Response: 396985860
Conc: 98.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.663 min
Delta R.T.: 0.000 min
Response: 169340236
Conc: 96.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.355 min
Delta R.T.: -0.004 min
Response: 247760121
Conc: 95.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.774 min
Delta R.T.: -0.002 min
Response: 106824219
Conc: 91.24 ng/ml



#3 AR-1016-1


R.T.: 5.654 min
Delta R.T.: -0.001 min
Response: 112470067
Conc: 952.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.768 min
Delta R.T.: 0.000 min
Response: 44933361
Conc: 914.68 ng/ml




#4 AR-1016-2

R.T.: 5.677 min
Delta R.T.: -0.001 min
Response: 163783972
Conc: 960.11 ng/ml



#4 AR-1016-2


R.T.: 4.787 min
Delta R.T.: 0.000 min
Response: 67872085
Conc: 929.87 ng/ml



#5 AR-1016-3


R.T.: 5.741 min
Delta R.T.: 0.000 min
Response: 100479703
Conc: 946.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 35529637
Conc: 912.20 ng/ml




#6 AR-1016-4

R.T.: 5.840 min
Delta R.T.: 0.000 min
Response: 81241696
Conc: 957.88 ng/ml



#6 AR-1016-4


R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 29433704
Conc: 890.78 ng/ml



#7 AR-1016-5


R.T.: 6.137 min
Delta R.T.: -0.001 min
Response: 78251058
Conc: 945.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 38046245
Conc: 907.89 ng/ml




#31 AR-1260-1

R.T.: 7.272 min
Delta R.T.: -0.001 min
Response: 141547931
Conc: 951.50 ng/ml



#31 AR-1260-1


R.T.: 6.276 min
Delta R.T.: 0.000 min
Response: 69356360
Conc: 915.45 ng/ml



#32 AR-1260-2


R.T.: 7.530 min
Delta R.T.: -0.002 min
Response: 162238735
Conc: 952.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.466 min
Delta R.T.: 0.000 min
Response: 80402414
Conc: 920.36 ng/ml




#33 AR-1260-3

R.T.: 7.891 min
Delta R.T.: -0.002 min
Response: 119948519
Conc: 956.24 ng/ml



#33 AR-1260-3


R.T.: 6.622 min
Delta R.T.: 0.000 min
Response: 75910361
Conc: 919.86 ng/ml



#34 AR-1260-4


R.T.: 8.120 min
Delta R.T.: -0.001 min
Response: 135822110
Conc: 962.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.101 min
Delta R.T.: 0.000 min
Response: 60168947
Conc: 916.46 ng/ml



#35 AR-1260-5

R.T.: 8.452 min
Delta R.T.: 0.000 min
Response: 269569227
Conc: 975.48 ng/ml



#35 AR-1260-5

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 126238253
Conc: 940.71 ng/ml