

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091621\
 Data File : P0081303.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Sep 2021 14:50
 Operator : AJ\MA
 Sample : M3759-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 20859

Manual Integrations
 APPROVED

mohammad
 9/17/2021 2:54:40 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 01:30:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 03 07:15:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.932	4.061	949312	369152	12.025m	14.938m
2) SA Decachlor...	10.989	9.412	1316238	515936	22.168	22.101

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091621\
Data File : PO081303.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Sep 2021 14:50
Operator : AJ\MA
Sample : M3759-03
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
20859


Manual Integrations
APPROVED

mohammad
9/17/2021 2:54:40 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 01:30:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 03 07:15:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.932 min

Delta R.T.: 0.002 min


Response: 949312

Conc: 12.03 ng/ml m

Instrument :
ECD_O
ClientSampleId :
20859

Manual Integrations
APPROVED

mohammad
9/17/2021 2:54:40 PM




#1 Tetrachloro-m-xylene

R.T.: 4.061 min

Delta R.T.: 0.005 min

Response: 369152

Conc: 14.94 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.989 min

Delta R.T.: -0.010 min

Response: 1316238

Conc: 22.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.412 min

Delta R.T.: -0.008 min

Response: 515936

Conc: 22.10 ng/ml