

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091622\
 Data File : P0089626.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Sep 2022 07:33
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 00:57:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 16 08:48:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.472	3.659	261.1E6	96278716	67.452	56.495
2) SA Decachlor...	10.345	8.762	177.6E6	62654512	71.945	55.525
Target Compounds						
26) L6 AR-1254-1	6.513	5.577	102.4E6	46145519	780.644	538.699 #
27) L6 AR-1254-2	6.733	5.726	130.8E6	39022464	660.833	522.718
28) L6 AR-1254-3	7.102	6.136	142.2E6	61754590	703.089	532.699
29) L6 AR-1254-4	7.389	6.367	99037525	37261391	719.872	577.892
30) L6 AR-1254-5	7.810	6.791	102.3E6	52410502	690.348	554.327



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091622\
Data File : PO089626.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Sep 2022 07:33
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 00:57:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 16 08:48:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.472 min
Delta R.T.: -0.004 min
Response: 261052385
Conc: 67.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.659 min
Delta R.T.: -0.004 min
Response: 96278716
Conc: 56.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.345 min
Delta R.T.: -0.014 min
Response: 177612146
Conc: 71.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.762 min
Delta R.T.: -0.015 min
Response: 62654512
Conc: 55.52 ng/ml



#26 AR-1254-1


R.T.: 6.513 min
Delta R.T.: -0.004 min
Response: 102402605
Conc: 780.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.577 min
Delta R.T.: -0.007 min
Response: 46145519
Conc: 538.70 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: -0.005 min
Response: 130849658
Conc: 660.83 ng/ml



#27 AR-1254-2


R.T.: 5.726 min
Delta R.T.: -0.007 min
Response: 39022464
Conc: 522.72 ng/ml



#28 AR-1254-3


R.T.: 7.102 min
Delta R.T.: -0.004 min
Response: 142164390
Conc: 703.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.136 min
Delta R.T.: -0.007 min
Response: 61754590
Conc: 532.70 ng/ml




#29 AR-1254-4

R.T.: 7.389 min
Delta R.T.: -0.004 min
Response: 99037525
Conc: 719.87 ng/ml



#29 AR-1254-4


R.T.: 6.367 min
Delta R.T.: -0.007 min
Response: 37261391
Conc: 577.89 ng/ml



#30 AR-1254-5

R.T.: 7.810 min
Delta R.T.: -0.005 min
Response: 102257108
Conc: 690.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.791 min
Delta R.T.: -0.007 min
Response: 52410502
Conc: 554.33 ng/ml