

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091719\
 Data File : P0060986.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Sep 2019 10:08
 Operator : SM/AJ
 Sample : K4825-09
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CSA-4-2

Manual Integrations
 APPROVED

Ankita
 9/19/2019 9:01:51 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 13:51:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.371	3.533	948836	753998	25.790	27.004
2)	SA Decachlor...	10.031	8.411	829090	530762	18.045	15.142
Target Compounds							
26)	L6 AR-1254-1	6.382	5.366	146537	146419	76.058m	71.045
27)	L6 AR-1254-2	6.602	5.509	234730	82942	74.590	43.954 #
28)	L6 AR-1254-3	6.968	5.904	375994	230545	106.757	72.510 #
29)	L6 AR-1254-4	7.250	6.128	115812	55322	42.781	26.572 #
30)	L6 AR-1254-5	7.667	6.539	301698	244314	102.311	77.195
31)	L7 AR-1260-1	7.128	6.032	155828	149724	64.005	71.182
32)	L7 AR-1260-2	7.378	6.217	428167	135107	141.620	52.377m#
33)	L7 AR-1260-3	7.741	6.367	104761	132150	43.898	54.878 #
34)	L7 AR-1260-4	7.962	6.832	137612	63099	49.188m	29.482m#
35)	L7 AR-1260-5	8.280	7.070	192750	161335	35.397	33.386m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091719\
Data File : PO060986.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Sep 2019 10:08
Operator : SM/AJ
Sample : K4825-09
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 13:51:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.371 min
Delta R.T.: 0.000 min
Response: 948836
Conc: 25.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM




#1 Tetrachloro-m-xylene

R.T.: 3.533 min
Delta R.T.: 0.001 min
Response: 753998
Conc: 27.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.031 min
Delta R.T.: 0.003 min
Response: 829090
Conc: 18.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.411 min
Delta R.T.: -0.002 min
Response: 530762
Conc: 15.14 ng/ml




#26 AR-1254-1

R.T.: 6.382 min
Delta R.T.: 0.000 min
Response: 146537
Conc: 76.06 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM




#26 AR-1254-1

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 146419
Conc: 71.05 ng/ml




#27 AR-1254-2

R.T.: 6.602 min
Delta R.T.: 0.002 min
Response: 234730
Conc: 74.59 ng/ml



#27 AR-1254-2

R.T.: 5.509 min
Delta R.T.: 0.000 min
Response: 82942
Conc: 43.95 ng/ml




#28 AR-1254-3

R.T.: 6.968 min
Delta R.T.: 0.003 min
Response: 375994
Conc: 106.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM




#28 AR-1254-3

R.T.: 5.904 min
Delta R.T.: 0.000 min
Response: 230545
Conc: 72.51 ng/ml




#29 AR-1254-4

R.T.: 7.250 min
Delta R.T.: 0.000 min
Response: 115812
Conc: 42.78 ng/ml



#29 AR-1254-4

R.T.: 6.128 min
Delta R.T.: 0.000 min
Response: 55322
Conc: 26.57 ng/ml




#30 AR-1254-5

R.T.: 7.667 min
Delta R.T.: 0.000 min
Response: 301698
Conc: 102.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM




#30 AR-1254-5

R.T.: 6.539 min
Delta R.T.: -0.001 min
Response: 244314
Conc: 77.19 ng/ml




#31 AR-1260-1

R.T.: 7.128 min
Delta R.T.: 0.000 min
Response: 155828
Conc: 64.01 ng/ml



#31 AR-1260-1

R.T.: 6.032 min
Delta R.T.: -0.003 min
Response: 149724
Conc: 71.18 ng/ml




#32 AR-1260-2

R.T.: 7.378 min
Delta R.T.: -0.006 min
Response: 428167
Conc: 141.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM




#32 AR-1260-2

R.T.: 6.217 min
Delta R.T.: -0.003 min
Response: 135107
Conc: 52.38 ng/ml m




#33 AR-1260-3

R.T.: 7.741 min
Delta R.T.: 0.000 min
Response: 104761
Conc: 43.90 ng/ml



#33 AR-1260-3

R.T.: 6.367 min
Delta R.T.: -0.003 min
Response: 132150
Conc: 54.88 ng/ml




#34 AR-1260-4

R.T.: 7.962 min
Delta R.T.: -0.003 min
Response: 137612
Conc: 49.19 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CSA-4-2


Manual Integrations
APPROVED

Ankita
9/19/2019 9:01:51 AM




#34 AR-1260-4

R.T.: 6.832 min
Delta R.T.: -0.003 min
Response: 63099
Conc: 29.48 ng/ml m



#35 AR-1260-5

R.T.: 8.280 min
Delta R.T.: 0.000 min
Response: 192750
Conc: 35.40 ng/ml



#35 AR-1260-5

R.T.: 7.070 min
Delta R.T.: -0.003 min
Response: 161335
Conc: 33.39 ng/ml m