

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091819\
 Data File : P0061063.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Sep 2019 10:38
 Operator : SM/AJ
 Sample : K4862-10
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CSA-4-5

Manual Integrations
 APPROVED

Ankita
 9/19/2019 4:37:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 12:44:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.370	3.531	897007	793164	24.381	28.406
2)	SA Decachlor...	10.029	8.409	862529	528963	18.773	15.091
Target Compounds							
26)	L6 AR-1254-1	6.383	5.363	223169	49348	115.833	23.945m#
27)	L6 AR-1254-2	6.601	5.508	75109	33001	23.868	17.488 #
28)	L6 AR-1254-3	6.968	5.902	127186	73445	36.112	23.099m#
29)	L6 AR-1254-4	7.248	6.127	33729	28985	12.460m	13.922
30)	L6 AR-1254-5	7.665	6.537	104186	118323	35.332m	37.386
31)	L7 AR-1260-1	7.127	6.030	51839	58421	21.293	27.775 #
32)	L7 AR-1260-2	7.379	6.215	127202	76490	42.073	29.653 #
33)	L7 AR-1260-3	7.740	6.364	53859	57192	22.569	23.750
34)	L7 AR-1260-4	7.961	6.830	79500	43089	28.416m	20.133 #
35)	L7 AR-1260-5	8.279	7.067	80791	80560	14.837	16.671m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091819\
Data File : PO061063.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Sep 2019 10:38
Operator : SM/AJ
Sample : K4862-10
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 12:44:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.370 min
Delta R.T.: 0.000 min
Response: 897007
Conc: 24.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM




#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: 0.000 min
Response: 793164
Conc: 28.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.029 min
Delta R.T.: 0.002 min
Response: 862529
Conc: 18.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.409 min
Delta R.T.: -0.004 min
Response: 528963
Conc: 15.09 ng/ml




#26 AR-1254-1

R.T.: 6.383 min
Delta R.T.: 0.002 min
Response: 223169
Conc: 115.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM




#26 AR-1254-1

R.T.: 5.363 min
Delta R.T.: -0.003 min
Response: 49348
Conc: 23.94 ng/ml m




#27 AR-1254-2

R.T.: 6.601 min
Delta R.T.: 0.000 min
Response: 75109
Conc: 23.87 ng/ml



#27 AR-1254-2

R.T.: 5.508 min
Delta R.T.: -0.002 min
Response: 33001
Conc: 17.49 ng/ml




#28 AR-1254-3

R.T.: 6.968 min
Delta R.T.: 0.002 min
Response: 127186
Conc: 36.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM




#28 AR-1254-3

R.T.: 5.902 min
Delta R.T.: -0.003 min
Response: 73445
Conc: 23.10 ng/ml m




#29 AR-1254-4

R.T.: 7.248 min
Delta R.T.: -0.002 min
Response: 33729
Conc: 12.46 ng/ml m



#29 AR-1254-4

R.T.: 6.127 min
Delta R.T.: -0.002 min
Response: 28985
Conc: 13.92 ng/ml




#30 AR-1254-5

R.T.: 7.665 min
Delta R.T.: -0.002 min
Response: 104186
Conc: 35.33 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM




#30 AR-1254-5

R.T.: 6.537 min
Delta R.T.: -0.003 min
Response: 118323
Conc: 37.39 ng/ml




#31 AR-1260-1

R.T.: 7.127 min
Delta R.T.: 0.000 min
Response: 51839
Conc: 21.29 ng/ml



#31 AR-1260-1

R.T.: 6.030 min
Delta R.T.: -0.005 min
Response: 58421
Conc: 27.77 ng/ml




#32 AR-1260-2

R.T.: 7.379 min
Delta R.T.: -0.005 min
Response: 127202
Conc: 42.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM




#32 AR-1260-2

R.T.: 6.215 min
Delta R.T.: -0.004 min
Response: 76490
Conc: 29.65 ng/ml




#33 AR-1260-3

R.T.: 7.740 min
Delta R.T.: 0.000 min
Response: 53859
Conc: 22.57 ng/ml



#33 AR-1260-3

R.T.: 6.364 min
Delta R.T.: -0.005 min
Response: 57192
Conc: 23.75 ng/ml




#34 AR-1260-4

R.T.: 7.961 min
Delta R.T.: -0.005 min
Response: 79500
Conc: 28.42 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CSA-4-5


Manual Integrations
APPROVED

Ankita
9/19/2019 4:37:46 PM




#34 AR-1260-4

R.T.: 6.830 min
Delta R.T.: -0.005 min
Response: 43089
Conc: 20.13 ng/ml



#35 AR-1260-5

R.T.: 8.279 min
Delta R.T.: 0.000 min
Response: 80791
Conc: 14.84 ng/ml



#35 AR-1260-5

R.T.: 7.067 min
Delta R.T.: -0.006 min
Response: 80560
Conc: 16.67 ng/ml m