

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091819\
 Data File : P0061101.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Sep 2019 22:05
 Operator : SM/AJ
 Sample : K4942-01
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-1

Manual Integrations
 APPROVED

Ankita
 9/19/2019 4:38:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 01:16:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.362	3.525	976898	791094	26.553	28.332m
2)	SA Decachlor...	10.020	8.401	819808	592802	17.843	16.912m
Target Compounds							
26)	L6 AR-1254-1	6.377	5.360	84032	91764	43.616	44.526
27)	L6 AR-1254-2	6.593	5.503	225877	49067	71.777	26.002 #
28)	L6 AR-1254-3	6.965	5.897	170684	73899	48.463	23.242 #
29)	L6 AR-1254-4	7.242	6.121	44940	36515	16.601m	17.539
30)	L6 AR-1254-5	7.659	6.531	121323	115003	41.143m	36.337
31)	L7 AR-1260-1	7.121	6.024	84435	83737	34.681	39.810
32)	L7 AR-1260-2	7.370	6.210	355898	76732	117.717	29.747 #
33)	L7 AR-1260-3	7.734	6.359	51034	66503	21.385m	27.617 #
34)	L7 AR-1260-4	7.957	6.824	83108	49467	29.706	23.113
35)	L7 AR-1260-5	8.273	7.063	88086	130045	16.176	26.911 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091819\
Data File : PO061101.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Sep 2019 22:05
Operator : SM/AJ
Sample : K4942-01
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 01:16:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.362 min
Delta R.T.: -0.008 min
Response: 976898
Conc: 26.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM




#1 Tetrachloro-m-xylene

R.T.: 3.525 min
Delta R.T.: -0.007 min
Response: 791094
Conc: 28.33 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.020 min
Delta R.T.: -0.007 min
Response: 819808
Conc: 17.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.401 min
Delta R.T.: -0.012 min
Response: 592802
Conc: 16.91 ng/ml m




#26 AR-1254-1

R.T.: 6.377 min
Delta R.T.: -0.005 min
Response: 84032
Conc: 43.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM




#26 AR-1254-1

R.T.: 5.360 min
Delta R.T.: -0.006 min
Response: 91764
Conc: 44.53 ng/ml




#27 AR-1254-2

R.T.: 6.593 min
Delta R.T.: -0.007 min
Response: 225877
Conc: 71.78 ng/ml



#27 AR-1254-2

R.T.: 5.503 min
Delta R.T.: -0.008 min
Response: 49067
Conc: 26.00 ng/ml




#28 AR-1254-3

R.T.: 6.965 min
Delta R.T.: 0.000 min
Response: 170684
Conc: 48.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM




#28 AR-1254-3

R.T.: 5.897 min
Delta R.T.: -0.007 min
Response: 73899
Conc: 23.24 ng/ml




#29 AR-1254-4

R.T.: 7.242 min
Delta R.T.: -0.008 min
Response: 44940
Conc: 16.60 ng/ml m



#29 AR-1254-4

R.T.: 6.121 min
Delta R.T.: -0.008 min
Response: 36515
Conc: 17.54 ng/ml




#30 AR-1254-5

R.T.: 7.659 min
Delta R.T.: -0.007 min
Response: 121323
Conc: 41.14 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM




#30 AR-1254-5

R.T.: 6.531 min
Delta R.T.: -0.009 min
Response: 115003
Conc: 36.34 ng/ml




#31 AR-1260-1

R.T.: 7.121 min
Delta R.T.: -0.007 min
Response: 84435
Conc: 34.68 ng/ml



#31 AR-1260-1

R.T.: 6.024 min
Delta R.T.: -0.010 min
Response: 83737
Conc: 39.81 ng/ml




#32 AR-1260-2

R.T.: 7.370 min
Delta R.T.: -0.014 min
Response: 355898
Conc: 117.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM




#32 AR-1260-2

R.T.: 6.210 min
Delta R.T.: -0.009 min
Response: 76732
Conc: 29.75 ng/ml




#33 AR-1260-3

R.T.: 7.734 min
Delta R.T.: -0.007 min
Response: 51034
Conc: 21.38 ng/ml m



#33 AR-1260-3

R.T.: 6.359 min
Delta R.T.: -0.010 min
Response: 66503
Conc: 27.62 ng/ml




#34 AR-1260-4

R.T.: 7.957 min
Delta R.T.: -0.008 min
Response: 83108
Conc: 29.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

Ankita
9/19/2019 4:38:26 PM




#34 AR-1260-4

R.T.: 6.824 min
Delta R.T.: -0.011 min
Response: 49467
Conc: 23.11 ng/ml



#35 AR-1260-5

R.T.: 8.273 min
Delta R.T.: -0.006 min
Response: 88086
Conc: 16.18 ng/ml



#35 AR-1260-5

R.T.: 7.063 min
Delta R.T.: -0.010 min
Response: 130045
Conc: 26.91 ng/ml