

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091820\
 Data File : P0071431.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Sep 2020 8:48
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 9/21/2020 1:04:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 11:44:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0090920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 09 16:04:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.343	3.526	3848584	1895059	54.307	48.540
2)	SA Decachlor...	9.959	8.707	5220790	2825932	55.984	51.302
Target Compounds							
3)	L1 AR-1016-1	5.647	4.750	1644293	835116	565.669	486.486
4)	L1 AR-1016-2	5.669	4.768	2423028	1176850	567.819	485.779
5)	L1 AR-1016-3	5.732	4.953	1429526	629700	563.543	496.116
6)	L1 AR-1016-4	5.839	5.010	1224845	551889	572.231	523.056
7)	L1 AR-1016-5	6.146	5.231	1215448	682374	593.810m	501.707
31)	L7 AR-1260-1	7.305	6.311	2596213	1417450	572.562	508.761
32)	L7 AR-1260-2	7.570	6.511	4055218	1732074	581.696	479.167
33)	L7 AR-1260-3	7.928	6.658	3396860	1493849	560.512	462.927
34)	L7 AR-1260-4	8.155	7.134	3174133	1371479	554.496	473.434
35)	L7 AR-1260-5	8.469	7.387	7319597	3574796	577.236	471.445

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091820\
Data File : PO071431.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Sep 2020 8:48
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 11:44:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO090920.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 09 16:04:54 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.343 min
Delta R.T.: -0.001 min
Response: 3848584
Conc: 54.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM




#1 Tetrachloro-m-xylene

R.T.: 3.526 min
Delta R.T.: -0.002 min
Response: 1895059
Conc: 48.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.959 min
Delta R.T.: 0.000 min
Response: 5220790
Conc: 55.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.707 min
Delta R.T.: -0.002 min
Response: 2825932
Conc: 51.30 ng/ml




#3 AR-1016-1

R.T.: 5.647 min
Delta R.T.: 0.000 min
Response: 1644293
Conc: 565.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM




#3 AR-1016-1

R.T.: 4.750 min
Delta R.T.: -0.002 min
Response: 835116
Conc: 486.49 ng/ml




#4 AR-1016-2

R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 2423028
Conc: 567.82 ng/ml



#4 AR-1016-2

R.T.: 4.768 min
Delta R.T.: -0.002 min
Response: 1176850
Conc: 485.78 ng/ml




#5 AR-1016-3

R.T.: 5.732 min
Delta R.T.: 0.000 min
Response: 1429526
Conc: 563.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM




#5 AR-1016-3

R.T.: 4.953 min
Delta R.T.: -0.003 min
Response: 629700
Conc: 496.12 ng/ml




#6 AR-1016-4

R.T.: 5.839 min
Delta R.T.: 0.000 min
Response: 1224845
Conc: 572.23 ng/ml



#6 AR-1016-4

R.T.: 5.010 min
Delta R.T.: -0.002 min
Response: 551889
Conc: 523.06 ng/ml




#7 AR-1016-5

R.T.: 6.146 min
Delta R.T.: 0.000 min
Response: 1215448
Conc: 593.81 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM




#7 AR-1016-5

R.T.: 5.231 min
Delta R.T.: -0.003 min
Response: 682374
Conc: 501.71 ng/ml




#31 AR-1260-1

R.T.: 7.305 min
Delta R.T.: 0.000 min
Response: 2596213
Conc: 572.56 ng/ml



#31 AR-1260-1

R.T.: 6.311 min
Delta R.T.: -0.002 min
Response: 1417450
Conc: 508.76 ng/ml




#32 AR-1260-2

R.T.: 7.570 min
Delta R.T.: 0.000 min
Response: 4055218
Conc: 581.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM




#32 AR-1260-2

R.T.: 6.511 min
Delta R.T.: -0.002 min
Response: 1732074
Conc: 479.17 ng/ml




#33 AR-1260-3

R.T.: 7.928 min
Delta R.T.: 0.000 min
Response: 3396860
Conc: 560.51 ng/ml



#33 AR-1260-3

R.T.: 6.658 min
Delta R.T.: -0.002 min
Response: 1493849
Conc: 462.93 ng/ml




#34 AR-1260-4

R.T.: 8.155 min
Delta R.T.: 0.000 min
Response: 3174133
Conc: 554.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
9/21/2020 1:04:59 PM




#34 AR-1260-4

R.T.: 7.134 min
Delta R.T.: -0.001 min
Response: 1371479
Conc: 473.43 ng/ml



#35 AR-1260-5

R.T.: 8.469 min
Delta R.T.: 0.000 min
Response: 7319597
Conc: 577.24 ng/ml



#35 AR-1260-5

R.T.: 7.387 min
Delta R.T.: -0.001 min
Response: 3574796
Conc: 471.44 ng/ml