

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0091823.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 18 17:23:15 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0098085.D 1000 =P0098083.D 750 =P0098084.D
 250 =P0098086.D 50 =P0098087.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.040	2.949	2.990	3.057	2.666	2.941	E6	5.42
2) SA	Decachlorobiphenyl	1.780	1.727	1.744	1.776	1.548	1.715	E6	5.61
3) L1	AR-1016-1	9.224	8.697	8.902	9.459	8.562	8.969	E4	4.13
4) L1	AR-1016-2	1.332	1.273	1.302	1.370	1.205	1.296	E5	4.82
5) L1	AR-1016-3	8.118	7.662	7.880	8.417	7.662	7.948	E4	4.06
6) L1	AR-1016-4	6.502	6.178	6.334	6.603	6.171	6.358	E4	3.03
7) L1	AR-1016-5	6.305	6.023	6.120	6.565	5.718	6.146	E4	5.14
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	1.123	1.061	1.085	1.159	1.040	1.094	E5	4.37
32) L7	AR-1260-2	1.291	1.235	1.255	1.320	1.180	1.256	E5	4.28
33) L7	AR-1260-3	9.168	8.853	9.035	9.294	7.941	8.858	E4	6.08
34) L7	AR-1260-4	1.039	0.995	1.018	1.047	0.873	0.994	E5	7.12
35) L7	AR-1260-5	2.084	2.025	2.041	2.105	1.876	2.026	E5	4.43
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0098085.D 1000 =P0098083.D 750 =P0098084.D
 250 =P0098086.D 50 =P0098087.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.473	1.421	1.441	1.507	1.420	1.453	E6	2.56
2) SA	Decachlorobiphenyl	0.976	0.929	0.946	1.070	0.990	0.982	E6	5.55
3) L1	AR-1016-1	5.246	4.825	4.976	5.589	5.521	5.232	E4	6.35
4) L1	AR-1016-2	7.130	6.730	6.891	7.471	7.264	7.097	E4	4.15
5) L1	AR-1016-3	3.898	3.612	3.735	4.158	3.972	3.875	E4	5.46

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0091823.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 18 17:23:15 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0098085.D 1000 =P0098083.D 750 =P0098084.D
 250 =P0098086.D 50 =P0098087.D

	Compound	500	1000	750	250	50	Avg		%RSD

6)	L1 AR-1016-4	3.132	2.852	2.970	3.417	3.359	3.146	E4	7.72
7)	L1 AR-1016-5	4.039	3.722	3.858	4.401	4.203	4.044	E4	6.67
8)	L2 AR-1221-1						0.000		-1.00
9)	L2 AR-1221-2						0.000		-1.00
10)	L2 AR-1221-3						0.000		-1.00
11)	L3 AR-1232-1						0.000		-1.00
12)	L3 AR-1232-2						0.000		-1.00
13)	L3 AR-1232-3						0.000		-1.00
14)	L3 AR-1232-4						0.000		-1.00
15)	L3 AR-1232-5						0.000		-1.00
16)	L4 AR-1242-1						0.000		-1.00
17)	L4 AR-1242-2						0.000		-1.00
18)	L4 AR-1242-3						0.000		-1.00
19)	L4 AR-1242-4						0.000		-1.00
20)	L4 AR-1242-5						0.000		-1.00
21)	L5 AR-1248-1						0.000		-1.00
22)	L5 AR-1248-2						0.000		-1.00
23)	L5 AR-1248-3						0.000		-1.00
24)	L5 AR-1248-4						0.000		-1.00
25)	L5 AR-1248-5						0.000		-1.00
26)	L6 AR-1254-1						0.000		-1.00
27)	L6 AR-1254-2						0.000		-1.00
28)	L6 AR-1254-3						0.000		-1.00
29)	L6 AR-1254-4						0.000		-1.00
30)	L6 AR-1254-5						0.000		-1.00
31)	L7 AR-1260-1	7.308	6.730	6.949	7.818	8.238	7.409	E4	8.37
32)	L7 AR-1260-2	8.550	7.905	8.137	9.163	9.397	8.630	E4	7.43
33)	L7 AR-1260-3	8.052	7.504	7.693	8.539	8.885	8.135	E4	7.08
34)	L7 AR-1260-4	6.420	5.991	6.166	6.869	6.991	6.487	E4	6.69
35)	L7 AR-1260-5	1.394	1.355	1.367	1.478	1.463	1.412	E5	3.98
36)	L8 AR-1262-1						0.000		-1.00
37)	L8 AR-1262-2						0.000		-1.00
38)	L8 AR-1262-3						0.000		-1.00
39)	L8 AR-1262-4						0.000		-1.00
40)	L8 AR-1262-5						0.000		-1.00
41)	L9 AR-1268-1						0.000		-1.00
42)	L9 AR-1268-2						0.000		-1.00
43)	L9 AR-1268-3						0.000		-1.00
44)	L9 AR-1268-4						0.000		-1.00
45)	L9 AR-1268-5						0.000		-1.00

(#) = Out of Range