

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091823\
 Data File : P0098089.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2023 17:41
 Operator : YP/AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 18:00:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Sep 18 18:00:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.501	3.614	148.5E6	71610588	50.000	50.000
2)	SA Decachlor...	10.350	8.630	85601092	46241030	50.000	50.000
Target Compounds							
8)	L2 AR-1221-1	4.711	3.829	12026620	6706411	328.882	337.790
9)	L2 AR-1221-2	4.799	3.914	10359354	6012416	385.007	390.800
10)	L2 AR-1221-3	4.876	3.991	32745555	18272231	408.267	417.412
11)	L3 AR-1232-1	4.876	3.991	32745555	18272231	500.000	500.000
12)	L3 AR-1232-2	5.410	4.719	16332683	16341349	500.000	500.000
13)	L3 AR-1232-3	5.704	4.895	29361000	8720788	500.000	500.000
14)	L3 AR-1232-4	5.866	4.979	13796469	8179442	500.000	500.000
15)	L3 AR-1232-5	6.003	5.151	11438350	8805991	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091823\
Data File : PO098089.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2023 17:41
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 18:00:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Sep 18 18:00:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.501 min
Delta R.T.: 0.000 min
Response: 148549526
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.614 min
Delta R.T.: 0.000 min
Response: 71610588
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.350 min
Delta R.T.: 0.000 min
Response: 85601092
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.630 min
Delta R.T.: 0.000 min
Response: 46241030
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.711 min
Delta R.T.: -0.002 min
Response: 12026620
Conc: 328.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 3.829 min
Delta R.T.: -0.001 min
Response: 6706411
Conc: 337.79 ng/ml




#9 AR-1221-2

R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 10359354
Conc: 385.01 ng/ml



#9 AR-1221-2


R.T.: 3.914 min
Delta R.T.: -0.001 min
Response: 6012416
Conc: 390.80 ng/ml



#10 AR-1221-3


R.T.: 4.876 min
Delta R.T.: 0.000 min
Response: 32745555
Conc: 408.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 3.991 min
Delta R.T.: 0.000 min
Response: 18272231
Conc: 417.41 ng/ml




#11 AR-1232-1

R.T.: 4.876 min
Delta R.T.: 0.000 min
Response: 32745555
Conc: 500.00 ng/ml



#11 AR-1232-1


R.T.: 3.991 min
Delta R.T.: 0.000 min
Response: 18272231
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 5.410 min
Delta R.T.: 0.000 min
Response: 16332683
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#12 AR-1232-2

R.T.: 4.719 min
Delta R.T.: 0.000 min
Response: 16341349
Conc: 500.00 ng/ml




#13 AR-1232-3

R.T.: 5.704 min
Delta R.T.: 0.000 min
Response: 29361000
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 8720788
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.866 min
Delta R.T.: 0.000 min
Response: 13796469
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#14 AR-1232-4

R.T.: 4.979 min
Delta R.T.: 0.000 min
Response: 8179442
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 6.003 min
Delta R.T.: 0.000 min
Response: 11438350
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.151 min
Delta R.T.: 0.000 min
Response: 8805991
Conc: 500.00 ng/ml