

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091823\
 Data File : P0098114.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 00:45
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 01:22:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 01:18:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.612	169.8E6	79443510	52.743	49.828
2) SA Decachlor...	10.348	8.635	98849533	47011182	53.565	48.828
Target Compounds						
26) L6 AR-1254-1	6.543	5.504	53666145	38044396	532.221	487.001
27) L6 AR-1254-2	6.763	5.653	80247329	33003527	530.092	486.144
28) L6 AR-1254-3	7.132	6.058	81939062	51524739	532.153	493.597
29) L6 AR-1254-4	7.418	6.287	58970285	31879671	532.136	496.855
30) L6 AR-1254-5	7.840	6.706	61646291	44433741	537.627	509.548



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091823\
Data File : PO098114.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2023 00:45
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 01:22:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 01:18:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.500 min
Delta R.T.: 0.000 min
Response: 169804848
Conc: 52.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.612 min
Delta R.T.: 0.000 min
Response: 79443510
Conc: 49.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: 0.000 min
Response: 98849533
Conc: 53.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.635 min
Delta R.T.: 0.003 min
Response: 47011182
Conc: 48.83 ng/ml



#26 AR-1254-1


R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 53666145
Conc: 532.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.504 min
Delta R.T.: 0.001 min
Response: 38044396
Conc: 487.00 ng/ml




#27 AR-1254-2

R.T.: 6.763 min
Delta R.T.: 0.000 min
Response: 80247329
Conc: 530.09 ng/ml



#27 AR-1254-2


R.T.: 5.653 min
Delta R.T.: 0.001 min
Response: 33003527
Conc: 486.14 ng/ml



#28 AR-1254-3


R.T.: 7.132 min
Delta R.T.: 0.000 min
Response: 81939062
Conc: 532.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.058 min
Delta R.T.: 0.002 min
Response: 51524739
Conc: 493.60 ng/ml




#29 AR-1254-4

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 58970285
Conc: 532.14 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: 0.002 min
Response: 31879671
Conc: 496.85 ng/ml



#30 AR-1254-5

R.T.: 7.840 min
Delta R.T.: 0.001 min
Response: 61646291
Conc: 537.63 ng/ml

Instrument :
ECD_O
ClientSampleId :



#30 AR-1254-5

R.T.: 6.706 min
Delta R.T.: 0.002 min
Response: 44433741
Conc: 509.55 ng/ml