

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091823\
 Data File : P0098125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 03:51
 Operator : YP/AJ
 Sample : 04413-03 10X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:49:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 02:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.614	4501498	2039120	1.531	1.404
2) SA Decachlor...	10.350	8.642	3014661	1317881	1.758	1.342
Target Compounds						
3) L1 AR-1016-1	5.671	4.712	201017	73872	2.241	1.412 #
4) L1 AR-1016-2	5.711	4.733	128370	24274	0.990	0.342 #
5) L1 AR-1016-3	5.771	4.913	302090	48355	3.801	1.248 #
6) L1 AR-1016-4	5.849	4.954	1028920	189932	16.184	6.037 #
7) L1 AR-1016-5	6.177	5.170	295694	92872	4.811	2.296 #
8) L2 AR-1221-1	4.738	3.850	2366801	108623	64.723	5.471 #
9) L2 AR-1221-2	4.813	3.921	847226	82224	31.487	5.344 #
10) L2 AR-1221-3	4.894	4.009	291354	51305	3.633	1.172 #
11) L3 AR-1232-1	4.894	4.009	291354	51305	4.449	1.404 #
12) L3 AR-1232-2	5.413	4.712	53876	73872	1.649	2.260 #
13) L3 AR-1232-3	5.711	4.913	128370	48355	2.186	2.772 #
14) L3 AR-1232-4	5.849	4.978	1028920	62383	37.289	3.813 #
15) L3 AR-1232-5	5.995	5.170	485535	92872	21.224	5.273 #
16) L4 AR-1242-1	5.671	4.712	201017	73872	2.492	1.589 #
17) L4 AR-1242-2	5.711	4.712	128370	73872	1.120	1.175
18) L4 AR-1242-3	5.771	4.913	302090	48355	4.282	1.409 #
19) L4 AR-1242-4	5.849	4.978	1028920	62383	18.287	1.759 #
20) L4 AR-1242-5	6.587	5.514	48802	167075	0.822	4.020 #
21) L5 AR-1248-1	5.671	4.712	201017	73872	3.417	2.149 #
22) L5 AR-1248-2	5.958	4.954	51522	189932	0.622	4.088 #
23) L5 AR-1248-3	6.177	4.978	295694	62383	3.298	1.256 #
24) L5 AR-1248-4	6.577	5.170	99684	92872	1.023	1.633 #
25) L5 AR-1248-5	6.587	5.579	48802	368862	0.486	6.633 #
26) L6 AR-1254-1	6.543	5.514	216776	167075	2.199	2.120
27) L6 AR-1254-2	6.768	5.641	101781	60438	0.694	0.875 #
28) L6 AR-1254-3	7.143	6.064	405634	143896	2.678	1.368 #
29) L6 AR-1254-4	7.400	6.288	49873	161076	0.461	2.517 #
30) L6 AR-1254-5	7.853	6.700	544712	50923	4.917	0.583 #
31) L7 AR-1260-1	7.282	6.195	80213	30532	0.733	0.412 #
32) L7 AR-1260-2	7.547	6.387	262723	127121	2.091	1.473 #
33) L7 AR-1260-3	7.923	6.533	190942	56553	2.155	0.695 #
34) L7 AR-1260-4	0.000	7.048f	0	32323	N.D.	0.498 #
35) L7 AR-1260-5	8.471	7.241	267773	67225	1.322	0.476 #
36) L8 AR-1262-1	7.923	6.814	190942	136524	1.397	3.765 #
37) L8 AR-1262-2	8.471	7.048f	267773	32323	1.124	0.430 #
38) L8 AR-1262-3	8.793	7.539	155330	1646445	0.984	27.764 #
39) L8 AR-1262-4	8.884	7.598	79042	29099	1.055	0.281 #
40) L8 AR-1262-5	9.564	0.000	93619	0	1.169	N.D. #
41) L9 AR-1268-1	8.793	7.539	155330	1646445	0.566	9.418 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091823\
 Data File : P0098125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 03:51
 Operator : YP/AJ
 Sample : 04413-03 10X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:49:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 02:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.884	7.598	79042	29099	0.322	0.196 #
43) L9	AR-1268-3	9.127	7.887	75999	212964	0.357	6.181 #
44) L9	AR-1268-4	9.564	0.000	93619	0	1.035	N.D. #
45) L9	AR-1268-5	9.991	8.382	396248	81654	0.570	0.228 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091823\
Data File : PO098125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2023 03:51
Operator : YP/AJ
Sample : 04413-03 10X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:49:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 02:14:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.500 min
Delta R.T.: 0.000 min
Response: 4501498
Conc: 1.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.614 min
Delta R.T.: -0.005 min
Response: 2039120
Conc: 1.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.350 min
Delta R.T.: -0.004 min
Response: 3014661
Conc: 1.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.642 min
Delta R.T.: 0.005 min
Response: 1317881
Conc: 1.34 ng/ml



#3 AR-1016-1


R.T.: 5.671 min
Delta R.T.: -0.010 min
Response: 201017
Conc: 2.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.712 min
Delta R.T.: 0.005 min
Response: 73872
Conc: 1.41 ng/ml




#4 AR-1016-2

R.T.: 5.711 min
Delta R.T.: 0.007 min
Response: 128370
Conc: 0.99 ng/ml



#4 AR-1016-2


R.T.: 4.733 min
Delta R.T.: 0.007 min
Response: 24274
Conc: 0.34 ng/ml



#5 AR-1016-3


R.T.: 5.771 min
Delta R.T.: 0.004 min
Response: 302090
Conc: 3.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.913 min
Delta R.T.: 0.012 min
Response: 48355
Conc: 1.25 ng/ml




#6 AR-1016-4

R.T.: 5.849 min
Delta R.T.: -0.017 min
Response: 1028920
Conc: 16.18 ng/ml



#6 AR-1016-4


R.T.: 4.954 min
Delta R.T.: 0.010 min
Response: 189932
Conc: 6.04 ng/ml



#7 AR-1016-5


R.T.: 6.177 min
Delta R.T.: 0.014 min
Response: 295694
Conc: 4.81 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.170 min
Delta R.T.: 0.013 min
Response: 92872
Conc: 2.30 ng/ml




#8 AR-1221-1

R.T.: 4.738 min
Delta R.T.: 0.025 min
Response: 2366801
Conc: 64.72 ng/ml



#8 AR-1221-1


R.T.: 3.850 min
Delta R.T.: 0.019 min
Response: 108623
Conc: 5.47 ng/ml



#9 AR-1221-2


R.T.: 4.813 min
Delta R.T.: 0.014 min
Response: 847226
Conc: 31.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.921 min
Delta R.T.: 0.006 min
Response: 82224
Conc: 5.34 ng/ml




#10 AR-1221-3

R.T.: 4.894 min
Delta R.T.: 0.017 min
Response: 291354
Conc: 3.63 ng/ml



#10 AR-1221-3


R.T.: 4.009 min
Delta R.T.: 0.017 min
Response: 51305
Conc: 1.17 ng/ml



#11 AR-1232-1


R.T.: 4.894 min
Delta R.T.: 0.018 min
Response: 291354
Conc: 4.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.009 min
Delta R.T.: 0.018 min
Response: 51305
Conc: 1.40 ng/ml




#12 AR-1232-2

R.T.: 5.413 min
Delta R.T.: 0.002 min
Response: 53876
Conc: 1.65 ng/ml



#12 AR-1232-2


R.T.: 4.712 min
Delta R.T.: -0.007 min
Response: 73872
Conc: 2.26 ng/ml



#13 AR-1232-3


R.T.: 5.711 min
Delta R.T.: 0.007 min
Response: 128370
Conc: 2.19 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.913 min
Delta R.T.: 0.018 min
Response: 48355
Conc: 2.77 ng/ml




#14 AR-1232-4

R.T.: 5.849 min
Delta R.T.: -0.017 min
Response: 1028920
Conc: 37.29 ng/ml



#14 AR-1232-4


R.T.: 4.978 min
Delta R.T.: 0.000 min
Response: 62383
Conc: 3.81 ng/ml



#15 AR-1232-5


R.T.: 5.995 min
Delta R.T.: -0.009 min
Response: 485535
Conc: 21.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.170 min
Delta R.T.: 0.019 min
Response: 92872
Conc: 5.27 ng/ml




#16 AR-1242-1

R.T.: 5.671 min
Delta R.T.: -0.010 min
Response: 201017
Conc: 2.49 ng/ml



#16 AR-1242-1


R.T.: 4.712 min
Delta R.T.: 0.011 min
Response: 73872
Conc: 1.59 ng/ml



#17 AR-1242-2


R.T.: 5.711 min
Delta R.T.: 0.007 min
Response: 128370
Conc: 1.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.712 min
Delta R.T.: -0.008 min
Response: 73872
Conc: 1.17 ng/ml




#18 AR-1242-3

R.T.: 5.771 min
Delta R.T.: 0.004 min
Response: 302090
Conc: 4.28 ng/ml



#18 AR-1242-3


R.T.: 4.913 min
Delta R.T.: 0.017 min
Response: 48355
Conc: 1.41 ng/ml



#19 AR-1242-4


R.T.: 5.849 min
Delta R.T.: -0.018 min
Response: 1028920
Conc: 18.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.978 min
Delta R.T.: -0.002 min
Response: 62383
Conc: 1.76 ng/ml




#20 AR-1242-5

R.T.: 6.587 min
Delta R.T.: -0.023 min
Response: 48802
Conc: 0.82 ng/ml



#20 AR-1242-5


R.T.: 5.514 min
Delta R.T.: 0.010 min
Response: 167075
Conc: 4.02 ng/ml



#21 AR-1248-1


R.T.: 5.671 min
Delta R.T.: -0.010 min
Response: 201017
Conc: 3.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.712 min
Delta R.T.: -0.004 min
Response: 73872
Conc: 2.15 ng/ml




#22 AR-1248-2

R.T.: 5.958 min
Delta R.T.: 0.000 min
Response: 51522
Conc: 0.62 ng/ml



#22 AR-1248-2


R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 189932
Conc: 4.09 ng/ml



#23 AR-1248-3


R.T.: 6.177 min
Delta R.T.: 0.014 min
Response: 295694
Conc: 3.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.978 min
Delta R.T.: -0.017 min
Response: 62383
Conc: 1.26 ng/ml




#24 AR-1248-4

R.T.: 6.577 min
Delta R.T.: 0.007 min
Response: 99684
Conc: 1.02 ng/ml



#24 AR-1248-4


R.T.: 5.170 min
Delta R.T.: 0.003 min
Response: 92872
Conc: 1.63 ng/ml



#25 AR-1248-5


R.T.: 6.587 min
Delta R.T.: -0.022 min
Response: 48802
Conc: 0.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.579 min
Delta R.T.: 0.018 min
Response: 368862
Conc: 6.63 ng/ml




#26 AR-1254-1

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 216776
Conc: 2.20 ng/ml



#26 AR-1254-1


R.T.: 5.514 min
Delta R.T.: 0.010 min
Response: 167075
Conc: 2.12 ng/ml



#27 AR-1254-2


R.T.: 6.768 min
Delta R.T.: 0.005 min
Response: 101781
Conc: 0.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.641 min
Delta R.T.: -0.011 min
Response: 60438
Conc: 0.87 ng/ml




#28 AR-1254-3

R.T.: 7.143 min
Delta R.T.: 0.011 min
Response: 405634
Conc: 2.68 ng/ml



#28 AR-1254-3


R.T.: 6.064 min
Delta R.T.: 0.008 min
Response: 143896
Conc: 1.37 ng/ml



#29 AR-1254-4


R.T.: 7.400 min
Delta R.T.: -0.019 min
Response: 49873
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.288 min
Delta R.T.: 0.002 min
Response: 161076
Conc: 2.52 ng/ml




#30 AR-1254-5

R.T.: 7.853 min
Delta R.T.: 0.014 min
Response: 544712
Conc: 4.92 ng/ml



#30 AR-1254-5


R.T.: 6.700 min
Delta R.T.: -0.004 min
Response: 50923
Conc: 0.58 ng/ml



#31 AR-1260-1


R.T.: 7.282 min
Delta R.T.: -0.016 min
Response: 80213
Conc: 0.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.195 min
Delta R.T.: 0.001 min
Response: 30532
Conc: 0.41 ng/ml




#32 AR-1260-2

R.T.: 7.547 min
Delta R.T.: -0.009 min
Response: 262723
Conc: 2.09 ng/ml



#32 AR-1260-2


R.T.: 6.387 min
Delta R.T.: 0.005 min
Response: 127121
Conc: 1.47 ng/ml



#33 AR-1260-3


R.T.: 7.923 min
Delta R.T.: 0.005 min
Response: 190942
Conc: 2.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.533 min
Delta R.T.: -0.003 min
Response: 56553
Conc: 0.70 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T.: 8.146 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 7.048 min
Delta R.T.: 0.040 min
Response: 32323
Conc: 0.50 ng/ml



#35 AR-1260-5


R.T.: 8.471 min
Delta R.T.: -0.004 min
Response: 267773
Conc: 1.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.241 min
Delta R.T.: -0.011 min
Response: 67225
Conc: 0.48 ng/ml




#36 AR-1262-1

R.T.: 7.923 min
Delta R.T.: 0.007 min
Response: 190942
Conc: 1.40 ng/ml



#36 AR-1262-1


R.T.: 6.814 min
Delta R.T.: 0.017 min
Response: 136524
Conc: 3.77 ng/ml



#37 AR-1262-2


R.T.: 8.471 min
Delta R.T.: -0.001 min
Response: 267773
Conc: 1.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.048 min
Delta R.T.: 0.045 min
Response: 32323
Conc: 0.43 ng/ml




#38 AR-1262-3

R.T.: 8.793 min
Delta R.T.: -0.002 min
Response: 155330
Conc: 0.98 ng/ml



#38 AR-1262-3


R.T.: 7.539 min
Delta R.T.: 0.009 min
Response: 1646445
Conc: 27.76 ng/ml



#39 AR-1262-4


R.T.: 8.884 min
Delta R.T.: 0.000 min
Response: 79042
Conc: 1.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.598 min
Delta R.T.: 0.004 min
Response: 29099
Conc: 0.28 ng/ml




#40 AR-1262-5

R.T.: 9.564 min
Delta R.T.: 0.000 min
Response: 93619
Conc: 1.17 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.090 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.793 min
Delta R.T.: 0.002 min
Response: 155330
Conc: 0.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.539 min
Delta R.T.: 0.008 min
Response: 1646445
Conc: 9.42 ng/ml




#42 AR-1268-2

R.T.: 8.884 min
Delta R.T.: -0.004 min
Response: 79042
Conc: 0.32 ng/ml



#42 AR-1268-2


R.T.: 7.598 min
Delta R.T.: 0.003 min
Response: 29099
Conc: 0.20 ng/ml



#43 AR-1268-3


R.T.: 9.127 min
Delta R.T.: 0.002 min
Response: 75999
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.887 min
Delta R.T.: -0.007 min
Response: 212964
Conc: 6.18 ng/ml




#44 AR-1268-4

R.T.: 9.564 min
Delta R.T.: 0.000 min
Response: 93619
Conc: 1.04 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.091 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.991 min
Delta R.T.: -0.003 min
Response: 396248
Conc: 0.57 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.382 min
Delta R.T.: 0.001 min
Response: 81654
Conc: 0.23 ng/ml