

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091823\
 Data File : P0098128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 04:42
 Operator : YP/AJ
 Sample : 04413-06 10X
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:48:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 02:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.615	4448542	2007180	1.513	1.382
2) SA Decachlor...	10.352	8.645	3323162	1089858	1.938	1.110 #
Target Compounds						
3) L1 AR-1016-1	5.690	4.660f	791849	118584	8.829	2.267 #
4) L1 AR-1016-2	5.690	4.786f	791849	34152	6.109	0.481 #
5) L1 AR-1016-3	5.756	4.896	2678003	68591	33.695	1.770 #
6) L1 AR-1016-4	5.827f	4.950	1344246	115116	21.143	3.659 #
7) L1 AR-1016-5	6.160	5.219f	192135	165362	3.126	4.089 #
8) L2 AR-1221-1	4.737	3.819	39322	33245	1.075	1.675 #
9) L2 AR-1221-2	4.815	3.895	522490	111388	19.418	7.240 #
10) L2 AR-1221-3	4.870	3.964	49356	348144	0.615	7.953 #
11) L3 AR-1232-1	4.870	3.964	49356	348144	0.754	9.527 #
12) L3 AR-1232-2	5.394	4.660f	54011	118584	1.653	3.628 #
13) L3 AR-1232-3	5.690	4.896	791849	68591	13.485	3.933 #
14) L3 AR-1232-4	5.827f	4.950	1344246	115116	48.717	7.037 #
15) L3 AR-1232-5	6.014	5.094f	128187	378312	5.603	21.480 #
16) L4 AR-1242-1	5.690	4.660f	791849	118584	9.817	2.551 #
17) L4 AR-1242-2	5.690	4.660f	791849	118584	6.910	1.885 #
18) L4 AR-1242-3	5.756	4.896	2678003	68591	37.963	1.998 #
19) L4 AR-1242-4	5.827f	4.950	1344246	115116	23.892	3.246 #
20) L4 AR-1242-5	6.630	5.526	632584	72957	10.654	1.756 #
21) L5 AR-1248-1	5.690	4.660f	791849	118584	13.462	3.449 #
22) L5 AR-1248-2	5.968	4.950	620522	115116	7.497	2.478 #
23) L5 AR-1248-3	6.160	5.023	192135	649037	2.143	13.070 #
24) L5 AR-1248-4	6.579	5.219f	162905	165362	1.671	2.907 #
25) L5 AR-1248-5	6.630	5.567	632584	62261	6.294	1.120 #
26) L6 AR-1254-1	6.539	5.526	91520	72957	0.928	0.926
27) L6 AR-1254-2	6.756	5.621	533822	29135	3.638	0.422 #
28) L6 AR-1254-3	7.130	6.058	166135	222003	1.097	2.110 #
29) L6 AR-1254-4	7.415	6.316	397684	63755	3.676	0.996 #
30) L6 AR-1254-5	7.844	6.704	521742	59992	4.710	0.687 #
31) L7 AR-1260-1	7.299	6.190	92723	178702	0.848	2.412 #
32) L7 AR-1260-2	7.562	6.365	604588	72240	4.812	0.837 #
33) L7 AR-1260-3	7.903	6.538	275437	33866	3.109	0.416 #
34) L7 AR-1260-4	8.138	7.008	509957	116282	5.129	1.792 #
35) L7 AR-1260-5	8.476	7.235	834511	133644	4.119	0.947 #
36) L8 AR-1262-1	7.903	6.760f	275437	14174	2.016	0.391 #
37) L8 AR-1262-2	8.476	7.008	834511	116282	3.501	1.546 #
38) L8 AR-1262-3	8.792	7.538	118958	751254	0.754	12.669 #
39) L8 AR-1262-4	8.867	7.538f	190658	751254	2.544	7.263 #
40) L8 AR-1262-5	9.553	8.084	192272	25840	2.402	0.539 #
41) L9 AR-1268-1	8.792	7.538	118958	751254	0.433	4.297 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091823\
 Data File : P0098128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 04:42
 Operator : YP/AJ
 Sample : 04413-06 10X
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:48:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 02:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.867	7.538f	190658	751254	0.778	5.064 #
43) L9	AR-1268-3	9.136	7.890	446187	38237	2.096	1.110 #
44) L9	AR-1268-4	9.553	8.084	192272	25840	2.126	0.475 #
45) L9	AR-1268-5	9.991	8.390	40883	32201	0.059	0.090 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091823\
 Data File : PO098128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 04:42
 Operator : YP/AJ
 Sample : 04413-06 10X
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:48:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 02:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.500 min
Delta R.T.: 0.000 min
Response: 4448542
Conc: 1.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.615 min
Delta R.T.: -0.005 min
Response: 2007180
Conc: 1.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.352 min
Delta R.T.: -0.003 min
Response: 3323162
Conc: 1.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: 0.008 min
Response: 1089858
Conc: 1.11 ng/ml



#3 AR-1016-1


R.T.: 5.690 min
Delta R.T.: 0.008 min
Response: 791849
Conc: 8.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.660 min
Delta R.T.: -0.046 min
Response: 118584
Conc: 2.27 ng/ml




#4 AR-1016-2

R.T.: 5.690 min
Delta R.T.: -0.014 min
Response: 791849
Conc: 6.11 ng/ml



#4 AR-1016-2


R.T.: 4.786 min
Delta R.T.: 0.060 min
Response: 34152
Conc: 0.48 ng/ml



#5 AR-1016-3


R.T.: 5.756 min
Delta R.T.: -0.010 min
Response: 2678003
Conc: 33.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.896 min
Delta R.T.: -0.005 min
Response: 68591
Conc: 1.77 ng/ml




#6 AR-1016-4

R.T.: 5.827 min
Delta R.T.: -0.039 min
Response: 1344246
Conc: 21.14 ng/ml



#6 AR-1016-4


R.T.: 4.950 min
Delta R.T.: 0.006 min
Response: 115116
Conc: 3.66 ng/ml



#7 AR-1016-5


R.T.: 6.160 min
Delta R.T.: -0.004 min
Response: 192135
Conc: 3.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.219 min
Delta R.T.: 0.062 min
Response: 165362
Conc: 4.09 ng/ml




#8 AR-1221-1

R.T.: 4.737 min
Delta R.T.: 0.024 min
Response: 39322
Conc: 1.08 ng/ml



#8 AR-1221-1


R.T.: 3.819 min
Delta R.T.: -0.012 min
Response: 33245
Conc: 1.67 ng/ml



#9 AR-1221-2


R.T.: 4.815 min
Delta R.T.: 0.016 min
Response: 522490
Conc: 19.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.895 min
Delta R.T.: -0.021 min
Response: 111388
Conc: 7.24 ng/ml




#10 AR-1221-3

R.T.: 4.870 min
Delta R.T.: -0.007 min
Response: 49356
Conc: 0.62 ng/ml



#10 AR-1221-3


R.T.: 3.964 min
Delta R.T.: -0.028 min
Response: 348144
Conc: 7.95 ng/ml



#11 AR-1232-1


R.T.: 4.870 min
Delta R.T.: -0.006 min
Response: 49356
Conc: 0.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.964 min
Delta R.T.: -0.027 min
Response: 348144
Conc: 9.53 ng/ml




#12 AR-1232-2

R.T.: 5.394 min
Delta R.T.: -0.016 min
Response: 54011
Conc: 1.65 ng/ml



#12 AR-1232-2


R.T.: 4.660 min
Delta R.T.: -0.059 min
Response: 118584
Conc: 3.63 ng/ml



#13 AR-1232-3


R.T.: 5.690 min
Delta R.T.: -0.014 min
Response: 791849
Conc: 13.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.896 min
Delta R.T.: 0.002 min
Response: 68591
Conc: 3.93 ng/ml




#14 AR-1232-4

R.T.: 5.827 min
Delta R.T.: -0.039 min
Response: 1344246
Conc: 48.72 ng/ml



#14 AR-1232-4


R.T.: 4.950 min
Delta R.T.: -0.029 min
Response: 115116
Conc: 7.04 ng/ml



#15 AR-1232-5


R.T.: 6.014 min
Delta R.T.: 0.010 min
Response: 128187
Conc: 5.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.094 min
Delta R.T.: -0.057 min
Response: 378312
Conc: 21.48 ng/ml




#16 AR-1242-1

R.T.: 5.690 min
Delta R.T.: 0.008 min
Response: 791849
Conc: 9.82 ng/ml



#16 AR-1242-1


R.T.: 4.660 min
Delta R.T.: -0.041 min
Response: 118584
Conc: 2.55 ng/ml



#17 AR-1242-2


R.T.: 5.690 min
Delta R.T.: -0.015 min
Response: 791849
Conc: 6.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.660 min
Delta R.T.: -0.060 min
Response: 118584
Conc: 1.89 ng/ml




#18 AR-1242-3

R.T.: 5.756 min
Delta R.T.: -0.011 min
Response: 2678003
Conc: 37.96 ng/ml



#18 AR-1242-3


R.T.: 4.896 min
Delta R.T.: 0.000 min
Response: 68591
Conc: 2.00 ng/ml



#19 AR-1242-4


R.T.: 5.827 min
Delta R.T.: -0.040 min
Response: 1344246
Conc: 23.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.950 min
Delta R.T.: -0.031 min
Response: 115116
Conc: 3.25 ng/ml




#20 AR-1242-5

R.T.: 6.630 min
Delta R.T.: 0.020 min
Response: 632584
Conc: 10.65 ng/ml



#20 AR-1242-5


R.T.: 5.526 min
Delta R.T.: 0.022 min
Response: 72957
Conc: 1.76 ng/ml



#21 AR-1248-1


R.T.: 5.690 min
Delta R.T.: 0.009 min
Response: 791849
Conc: 13.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.660 min
Delta R.T.: -0.056 min
Response: 118584
Conc: 3.45 ng/ml




#22 AR-1248-2

R.T.: 5.968 min
Delta R.T.: 0.010 min
Response: 620522
Conc: 7.50 ng/ml



#22 AR-1248-2


R.T.: 4.950 min
Delta R.T.: -0.004 min
Response: 115116
Conc: 2.48 ng/ml



#23 AR-1248-3


R.T.: 6.160 min
Delta R.T.: -0.003 min
Response: 192135
Conc: 2.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.023 min
Delta R.T.: 0.028 min
Response: 649037
Conc: 13.07 ng/ml




#24 AR-1248-4

R.T.: 6.579 min
Delta R.T.: 0.009 min
Response: 162905
Conc: 1.67 ng/ml



#24 AR-1248-4


R.T.: 5.219 min
Delta R.T.: 0.052 min
Response: 165362
Conc: 2.91 ng/ml



#25 AR-1248-5


R.T.: 6.630 min
Delta R.T.: 0.021 min
Response: 632584
Conc: 6.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.567 min
Delta R.T.: 0.006 min
Response: 62261
Conc: 1.12 ng/ml




#26 AR-1254-1

R.T.: 6.539 min
Delta R.T.: -0.004 min
Response: 91520
Conc: 0.93 ng/ml



#26 AR-1254-1


R.T.: 5.526 min
Delta R.T.: 0.022 min
Response: 72957
Conc: 0.93 ng/ml



#27 AR-1254-2


R.T.: 6.756 min
Delta R.T.: -0.007 min
Response: 533822
Conc: 3.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.621 min
Delta R.T.: -0.031 min
Response: 29135
Conc: 0.42 ng/ml




#28 AR-1254-3

R.T.: 7.130 min
Delta R.T.: -0.002 min
Response: 166135
Conc: 1.10 ng/ml



#28 AR-1254-3


R.T.: 6.058 min
Delta R.T.: 0.002 min
Response: 222003
Conc: 2.11 ng/ml



#29 AR-1254-4


R.T.: 7.415 min
Delta R.T.: -0.004 min
Response: 397684
Conc: 3.68 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.316 min
Delta R.T.: 0.030 min
Response: 63755
Conc: 1.00 ng/ml




#30 AR-1254-5

R.T.: 7.844 min
Delta R.T.: 0.005 min
Response: 521742
Conc: 4.71 ng/ml



#30 AR-1254-5


R.T.: 6.704 min
Delta R.T.: 0.000 min
Response: 59992
Conc: 0.69 ng/ml



#31 AR-1260-1


R.T.: 7.299 min
Delta R.T.: 0.000 min
Response: 92723
Conc: 0.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.190 min
Delta R.T.: -0.004 min
Response: 178702
Conc: 2.41 ng/ml




#32 AR-1260-2

R.T.: 7.562 min
Delta R.T.: 0.006 min
Response: 604588
Conc: 4.81 ng/ml



#32 AR-1260-2


R.T.: 6.365 min
Delta R.T.: -0.017 min
Response: 72240
Conc: 0.84 ng/ml



#33 AR-1260-3


R.T.: 7.903 min
Delta R.T.: -0.015 min
Response: 275437
Conc: 3.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.538 min
Delta R.T.: 0.003 min
Response: 33866
Conc: 0.42 ng/ml




#34 AR-1260-4

R.T.: 8.138 min
Delta R.T.: -0.008 min
Response: 509957
Conc: 5.13 ng/ml



#34 AR-1260-4


R.T.: 7.008 min
Delta R.T.: 0.000 min
Response: 116282
Conc: 1.79 ng/ml



#35 AR-1260-5


R.T.: 8.476 min
Delta R.T.: 0.001 min
Response: 834511
Conc: 4.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.235 min
Delta R.T.: -0.017 min
Response: 133644
Conc: 0.95 ng/ml




#36 AR-1262-1

R.T.: 7.903 min
Delta R.T.: -0.013 min
Response: 275437
Conc: 2.02 ng/ml



#36 AR-1262-1


R.T.: 6.760 min
Delta R.T.: -0.037 min
Response: 14174
Conc: 0.39 ng/ml



#37 AR-1262-2


R.T.: 8.476 min
Delta R.T.: 0.004 min
Response: 834511
Conc: 3.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.008 min
Delta R.T.: 0.004 min
Response: 116282
Conc: 1.55 ng/ml




#38 AR-1262-3

R.T.: 8.792 min
Delta R.T.: -0.003 min
Response: 118958
Conc: 0.75 ng/ml



#38 AR-1262-3


R.T.: 7.538 min
Delta R.T.: 0.008 min
Response: 751254
Conc: 12.67 ng/ml



#39 AR-1262-4


R.T.: 8.867 min
Delta R.T.: -0.018 min
Response: 190658
Conc: 2.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.538 min
Delta R.T.: -0.056 min
Response: 751254
Conc: 7.26 ng/ml




#40 AR-1262-5

R.T.: 9.553 min
Delta R.T.: -0.011 min
Response: 192272
Conc: 2.40 ng/ml



#40 AR-1262-5


R.T.: 8.084 min
Delta R.T.: -0.006 min
Response: 25840
Conc: 0.54 ng/ml



#41 AR-1268-1


R.T.: 8.792 min
Delta R.T.: 0.000 min
Response: 118958
Conc: 0.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.538 min
Delta R.T.: 0.007 min
Response: 751254
Conc: 4.30 ng/ml




#42 AR-1268-2

R.T.: 8.867 min
Delta R.T.: -0.022 min
Response: 190658
Conc: 0.78 ng/ml



#42 AR-1268-2


R.T.: 7.538 min
Delta R.T.: -0.057 min
Response: 751254
Conc: 5.06 ng/ml



#43 AR-1268-3


R.T.: 9.136 min
Delta R.T.: 0.011 min
Response: 446187
Conc: 2.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.890 min
Delta R.T.: -0.004 min
Response: 38237
Conc: 1.11 ng/ml




#44 AR-1268-4

R.T.: 9.553 min
Delta R.T.: -0.011 min
Response: 192272
Conc: 2.13 ng/ml



#44 AR-1268-4


R.T.: 8.084 min
Delta R.T.: -0.006 min
Response: 25840
Conc: 0.48 ng/ml



#45 AR-1268-5

R.T.: 9.991 min
Delta R.T.: -0.003 min
Response: 40883
Conc: 0.06 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.390 min
Delta R.T.: 0.010 min
Response: 32201
Conc: 0.09 ng/ml