

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091919\
 Data File : P0061159.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Sep 2019 20:25
 Operator : SM/AJ
 Sample : K4959-02
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BASE-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 01:27:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 14 00:06:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.366	3.530	840986	640208	22.859	22.928
2)	SA Decachlor...	10.025	8.405	930242	694803	20.246	19.822
Target Compounds							
31)	L7 AR-1260-1	7.125	6.029	77689	73657	31.910	35.018
32)	L7 AR-1260-2	7.381	6.214	98653	93663	32.630	36.311
33)	L7 AR-1260-3	7.738	6.364	64938	79920	27.211	33.189
34)	L7 AR-1260-4	7.964	6.828	83930	62206	30.000	29.065
35)	L7 AR-1260-5	8.277	7.067	132888	126076	24.404	26.089



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091919\
Data File : P0061159.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Sep 2019 20:25
Operator : SM/AJ
Sample : K4959-02
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
BASE-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 01:27:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091319.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 14 00:06:36 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.366 min
Delta R.T.: -0.004 min
Response: 840986
Conc: 22.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
BASE-2




#1 Tetrachloro-m-xylene

R.T.: 3.530 min
Delta R.T.: -0.001 min
Response: 640208
Conc: 22.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.025 min
Delta R.T.: -0.002 min
Response: 930242
Conc: 20.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.405 min
Delta R.T.: -0.008 min
Response: 694803
Conc: 19.82 ng/ml



#31 AR-1260-1


R.T.: 7.125 min
Delta R.T.: -0.003 min
Response: 77689
Conc: 31.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
BASE-2




#31 AR-1260-1

R.T.: 6.029 min
Delta R.T.: -0.006 min
Response: 73657
Conc: 35.02 ng/ml




#32 AR-1260-2

R.T.: 7.381 min
Delta R.T.: -0.003 min
Response: 98653
Conc: 32.63 ng/ml



#32 AR-1260-2


R.T.: 6.214 min
Delta R.T.: -0.006 min
Response: 93663
Conc: 36.31 ng/ml



#33 AR-1260-3


R.T.: 7.738 min
Delta R.T.: -0.003 min
Response: 64938
Conc: 27.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
BASE-2




#33 AR-1260-3

R.T.: 6.364 min
Delta R.T.: -0.006 min
Response: 79920
Conc: 33.19 ng/ml




#34 AR-1260-4

R.T.: 7.964 min
Delta R.T.: -0.002 min
Response: 83930
Conc: 30.00 ng/ml



#34 AR-1260-4


R.T.: 6.828 min
Delta R.T.: -0.007 min
Response: 62206
Conc: 29.06 ng/ml



#35 AR-1260-5

R.T.: 8.277 min
Delta R.T.: -0.002 min
Response: 132888
Conc: 24.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
BASE-2



#35 AR-1260-5

R.T.: 7.067 min
Delta R.T.: -0.006 min
Response: 126076
Conc: 26.09 ng/ml