

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091923\
 Data File : P0098156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 03:23
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 04:13:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 04:13:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.497	3.631	14478553	5828203	4.637	4.358
2)	SA Decachlor...	10.343	8.657	8528662	5252481	4.606	4.598
Target Compounds							
26)	L6 AR-1254-1	6.540	5.526	4710571	3679792	47.901	48.913
27)	L6 AR-1254-2	6.761	5.674	7183471	3505657	48.306	52.416
28)	L6 AR-1254-3	7.129	6.079	7251173	4942627	48.015	47.455
29)	L6 AR-1254-4	7.416	6.309	5112098	2949049	47.306	45.995
30)	L6 AR-1254-5	7.838	6.728	5205163	4401787	46.346	48.573



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091923\
Data File : PO098156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2023 03:23
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 04:13:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 04:13:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 14478553
Conc: 4.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254IC050




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 5828203
Conc: 4.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.343 min
Delta R.T.: 0.000 min
Response: 8528662
Conc: 4.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.657 min
Delta R.T.: 0.000 min
Response: 5252481
Conc: 4.60 ng/ml



#26 AR-1254-1


R.T.: 6.540 min
Delta R.T.: 0.000 min
Response: 4710571
Conc: 47.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254\ICC050




#26 AR-1254-1

R.T.: 5.526 min
Delta R.T.: 0.000 min
Response: 3679792
Conc: 48.91 ng/ml




#27 AR-1254-2

R.T.: 6.761 min
Delta R.T.: 0.000 min
Response: 7183471
Conc: 48.31 ng/ml



#27 AR-1254-2


R.T.: 5.674 min
Delta R.T.: 0.000 min
Response: 3505657
Conc: 52.42 ng/ml



#28 AR-1254-3


R.T.: 7.129 min
Delta R.T.: 0.000 min
Response: 7251173
Conc: 48.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 4942627
Conc: 47.46 ng/ml




#29 AR-1254-4

R.T.: 7.416 min
Delta R.T.: 0.000 min
Response: 5112098
Conc: 47.31 ng/ml



#29 AR-1254-4


R.T.: 6.309 min
Delta R.T.: 0.000 min
Response: 2949049
Conc: 46.00 ng/ml



#30 AR-1254-5

R.T.: 7.838 min
Delta R.T.: 0.000 min
Response: 5205163
Conc: 46.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.728 min
Delta R.T.: 0.000 min
Response: 4401787
Conc: 48.57 ng/ml