

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091923\
 Data File : P0098157.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 03:40
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/20/2023
 Supervised By :Ankita Jodhani 09/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 04:16:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 04:15:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.632	162.0E6	68372992	50.000	50.000
2) SA Decachlor...	10.345	8.658	97445060	57595834	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.915	6.767	71483678	45774368	500.000	500.000
37) L8 AR-1262-2	8.470	7.027	123.1E6	40743974	500.000	500.000
38) L8 AR-1262-3	8.793	7.554	82618175	31789924	500.000	500.000
39) L8 AR-1262-4	8.884	7.618	37901646	58289555	498.727m	500.000
40) L8 AR-1262-5	9.561	8.114	43073940	27152690	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091923\
Data File : PO098157.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2023 03:40
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2023
Supervised By :Ankita Jodhani 09/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 04:16:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 04:15:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 162002384
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2023
Supervised By :Ankita Jodhani 09/20/2023




#1 Tetrachloro-m-xylene

R.T.: 3.632 min
Delta R.T.: 0.000 min
Response: 68372992
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.345 min
Delta R.T.: 0.000 min
Response: 97445060
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.658 min
Delta R.T.: 0.000 min
Response: 57595834
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 7.915 min
Delta R.T.: 0.000 min
Response: 71483678
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2023
Supervised By :Ankita Jodhani 09/20/2023




#36 AR-1262-1

R.T.: 6.767 min
Delta R.T.: 0.000 min
Response: 45774368
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.470 min
Delta R.T.: 0.000 min
Response: 123118480
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.027 min
Delta R.T.: 0.000 min
Response: 40743974
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 8.793 min
Delta R.T.: 0.000 min
Response: 82618175
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2023
Supervised By :Ankita Jodhani 09/20/2023




#38 AR-1262-3

R.T.: 7.554 min
Delta R.T.: 0.000 min
Response: 31789924
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.884 min
Delta R.T.: 0.000 min
Response: 37901646
Conc: 498.73 ng/ml m



#39 AR-1262-4

R.T.: 7.618 min
Delta R.T.: 0.000 min
Response: 58289555
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 9.561 min
Delta R.T.: 0.000 min
Response: 43073940
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/20/2023
Supervised By :Ankita Jodhani 09/20/2023



#40 AR-1262-5

R.T.: 8.114 min
Delta R.T.: 0.000 min
Response: 27152690
Conc: 500.00 ng/ml