

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0091923\
 Data File : P0098159.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 04:13
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 05:28:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0091923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 20 05:28:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.497	3.633	238.0E6	102.3E6	75.090	75.109
2) SA Decachlor...	10.345	8.658	247.9E6	145.1E6	74.588	74.689
Target Compounds						
41) L9 AR-1268-1	8.789	7.554	211.9E6	135.3E6	749.521	746.199
42) L9 AR-1268-2	8.886	7.618	189.0E6	120.4E6	750.714	745.820
43) L9 AR-1268-3	9.121	7.916	163.6E6	25403913	749.838	744.479
44) L9 AR-1268-4	9.560	8.113	69954323	43466439	750.937	745.180
45) L9 AR-1268-5	9.991	8.403	549.2E6	322.6E6	749.229	749.729



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO091923\
Data File : PO098159.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2023 04:13
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 05:28:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO091923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 20 05:28:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.497 min
Delta R.T.: 0.000 min
Response: 237986585
Conc: 75.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: 0.000 min
Response: 102277494
Conc: 75.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.345 min
Delta R.T.: 0.000 min
Response: 247909526
Conc: 74.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.658 min
Delta R.T.: 0.000 min
Response: 145073479
Conc: 74.69 ng/ml



#41 AR-1268-1


R.T.: 8.789 min
Delta R.T.: 0.000 min
Response: 211929393
Conc: 749.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 7.554 min
Delta R.T.: 0.000 min
Response: 135306767
Conc: 746.20 ng/ml




#42 AR-1268-2

R.T.: 8.886 min
Delta R.T.: 0.000 min
Response: 189014382
Conc: 750.71 ng/ml



#42 AR-1268-2


R.T.: 7.618 min
Delta R.T.: 0.000 min
Response: 120438063
Conc: 745.82 ng/ml



#43 AR-1268-3


R.T.: 9.121 min
Delta R.T.: 0.000 min
Response: 163648299
Conc: 749.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 7.916 min
Delta R.T.: 0.000 min
Response: 25403913
Conc: 744.48 ng/ml




#44 AR-1268-4

R.T.: 9.560 min
Delta R.T.: 0.000 min
Response: 69954323
Conc: 750.94 ng/ml



#44 AR-1268-4


R.T.: 8.113 min
Delta R.T.: 0.000 min
Response: 43466439
Conc: 745.18 ng/ml



#45 AR-1268-5

R.T.: 9.991 min
Delta R.T.: 0.000 min
Response: 549172286
Conc: 749.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 8.403 min
Delta R.T.: 0.000 min
Response: 322557911
Conc: 749.73 ng/ml